

1	2	3	4	5	6	7	8	8	9
85	Digital Balance	<p>Full digital calibration, sharp, easy to read fluorescent / LCD display, sealed key pad protect against dusts and spills, round pan, calibration weight accurately to 05 decimal places with balance housing having door opening / closing configuration to suit both right and left handed operators, build in level and leveling feet, touch bottom operation, tare facility, capacity up to 180 gm., operating temp. 00-500C, optional RS 232C interface, measurement data can be sent to a printer or to a computer, power supply 220-240 V AC, 50 Hz, stabilization time approximately 2-5 Secs.</p> <p>Max. capacity : 120g Readability : 0.01 mg Repeatability : 0.04 mg Repeatability (sd) fine range : 0.02 mg Repeatability (sd) fine range at low load : 0.02 mg Linearity : 0.1 mg Linearity : 0.02 mg (Within 10gm) Sensitivity : 2 ppm/°C Stabilization time: 8s Weighing pan dimension: 80 mm Wt. of Balance: 6.6 kg TWT Weighing cell with two built-in weights for constant accuracy over the entire weighing range Fully automatic time and temp. controlled internal adjustment.</p>	SARTORIUS INDIA PVT LTD	SARTORIUS INDIA PVT LTD	SECUR A 125 10 IN	635,160.00	440,800.00	Kannala Agencies	L.I

22.08.2019

21/08/19
 22/08/2019
 22/08/2019

22/08/2019




1	2	3	4	5	6	7	8	8	9
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediocal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze needle, 2ml + 22 gauze needle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	10 ml + 18 Gauze needle	12.18	8.70	Kamala Agencies	L.I
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediocal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze needle, 2ml + 22 gauze needle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	20 ml + 18 Gauze needle	18.20	13.00	Kamala Agencies	L.I
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediocal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze needle, 2ml + 22 gauze needle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz 50ml + 18 gauze needle	JORGEN KRUUSE	KRUUSE	20ml + 18 gauze needle,	24.00	13.00	National Enterprises	L.I
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediocal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze needle, 2ml + 22 gauze needle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	5 ml + 20 Gauze needle	8.26	5.90	Kamala Agencies	L.I

21/04/19
 21/04/19
 21/04/19
 21/04/19

21/04/19
 21/04/19

21/04/19
 21/04/19

1	2	3	4	5	6	7	8	9
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediactal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze middle, 2ml + 22 gauze middle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz, 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	2 ml + 22 Gauze needle	6.00	4.10	Kamala Agencies
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediactal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze middle, 2ml + 22 gauze middle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz, 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	5 ml + 18 Gauze needle	8.26	5.90	Kamala Agencies
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediactal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze middle, 2ml + 22 gauze middle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz, 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	2 ml + 18 Gauze needle	6.00	4.10	Kamala Agencies
92	Disposable Syringe with needle	<u>Disposable syringe</u> Sterile single use disposable syringe ,flow wrap package using mediactal grade) - 10ml + 18 gauze needle, 20ml + 18 gauze needle, 5ml + 20 gauze middle, 2ml + 22 gauze middle, 5ml + 18 gauze needle, 2ml + 18 gauze needle, needle without syringe- 22gz,20gz,18gz,16gz, 50ml + 18 gauze needle	P.H. HEALTHCARE LTD	PARAS	50 ml + 18 Gauze needle	27.00	19.00	Kamala Agencies

1	2	3	4	5	6	7	8	9
93	Dipstick for LN2 Container	length will be 1 meter, made of plastic and folding type should have double graduation in Cms and inches	Newick Instruments	Newtek	Each	340.00	210.00	National Enterprises L.L
94	Dressing table for small animal	(SS) • Made of stainless steel Frame of 30mm Square x 18 G thickness Table top of stainless steel sheet of 202 Grade x 18 G in size 4 feet x 2.5 feet. • Total Height of Table is 3 feet. Top is having slope towards center with center hole having four grooves towards the drainage hole, to facilitate drainage of fluids from all areas of the table top. Table Frame is supported by cross Bar of 25 mm Square/round Pipe made of 202 Grade Stainless steel in 18G thickness. Rubber Stoppers are provided on all four legs of table.	NATIONAL MEDITEK	NATOMED	EACH	32,800.00	19,500.00	Kamala Agencies L.L
95	Disposable earloop face mask high Bacteria Filtration	Spun bond PP Fabric which ensures that these masks won't harm the skin if worn for longer period of time. It has strong tie band. High Bacteria Filtration (3-Ply: BFE: 99%) low Breathing resistance* (3-Ply: Delta P ~ 29, 4 Pa), SENSITIVE, INNER LAYER MADE OF SOFT FACIAL TISSUE, NO DYE, NO CHEMICAL, FLUID RESISTANT, LATEX FREE, FIBER-GLASS FREE, ODORLESS	NATIONAL MEDITEK	NATOMED	EACH	7.00	2.50	Kamala Agencies L.L

2/10/2019
 2/10/2019
 2/10/2019
 2/10/2019

2/10/2019
 2/10/2019

2/10/2019
 2/10/2019

1	2	3	4	5	6	7	8	8	9
96	Dispenser	<p>AUTO DISPENSER FOR SEMEN EXTENDER Programmable dispenser for measuring and dispensing the required quantity of extender based up on user defined protocol.</p> <ol style="list-style-type: none"> 1. The system should be with microprocessor based stepper motor. 2. The system should have range of the dispensing volume from 4 cc/stroke to 100 cc/ stroke. 3. Should have flow rate of 0.01 ml/min to 750 ml/min. 4. Should have step rate between 27.3 sec/step to 416.7 µsec/step. 5. Should have maximum Back Pressure of 30 PSI 6. The system should have option for RS232 Interface and choice of multiple Baud rate settings. 7. The system should be equipped with digital weighing balance with the accuracy 0.01 gm. 8. The system should be paired with weighing balance and the pump should stop automatically once the desired volume is dispensed. 	IMV INDIA PVT LTD	IMV, Made in France	EACH	2,016,000.00	1,260,000.00	Kamala Agencies	L1

20/10/19
 HHG
 20/10/19

22.08.2019
 22.08.2019

22.08.2019
 22.08.2019

1	2	3	4	5	6	7	8	9
		<p>9. The system should have option for reverse movement at the end of the cycle to prevent the fast drop, to maintain the accuracy of volume dispensed.</p> <p>10. The System should have option reverse rotation for emptying the Silicon Tube after use.</p> <p>11. The system should be compatible with Software for Management of Integrated Laboratory Environment</p> <p>The Equipment must be quoted one year warranty including onsite training & installation. The concern should have experience of installing successfully at least (02) two such systems in CMU Graded Semen Labs in India, for which necessary documents must be attached. The satisfactory working report from the user also to be attached. The supplier should have their services center in India to provide after sales services and all relevant details should be provided in tender.</p>						
97	Disposable Complete for semen collection	Contains disposable apron, caps and gloves	JORGEN KRUUSE	KRUUSE	Each	16,320.00	10,250.00	National Enterprises
98	Disposable sterile petridish	Polypropelene, Diameter 3inch, size-80x15mm-plain	Himedia Laboratories Pvt Ltd	Himedia	600 nos/Each	7,440.00	6,400.00	National Enterprises
99	DISTILLED WATER PLANT	<p>FEED WATER: tap water</p> <p>PRODUCT WATER QUALITY</p> <p>Flow Rate: upto 2L/min</p> <p>Resistivity: $\geq 2M\Omega \cdot cm$ (25°C)</p> <p>TOC: < 50 ppb</p> <p>Bacteria: < 1 cfu/ml</p> <p>Foot Print (HxWxD): 39x30x41cm.</p> <p>Operating Weight: 23 Kg</p>	P.L. TANDON & CO	TANCO	EACH	60,000.00	38,900.00	Kamala Agencies

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

1	2	3	4	5	6	7	8	8	9
100	Deep freezer Along with servo stabilizer (80 degree C)	Fully programmable microprocessor controlled VERTICAL type freezer with capacity of 400- 500 liters. 2. Freezer temperature should be up to - 80°C with 1°C increment at room temp. 3. Stage compressors working in a cascade for maximum heat removal and quick recovery to set temperature with at least 1°C increment at room temp. 4. CFC-FREE, HCFC-FREE non flammable refrigerant compressor should be capable to run on any voltage between 190-270 V. 5. Freezer should have at least five insulated doors giving access to shelves which should be adjustable in steps. must come with 4 adjustable shelves and make the floor cabinet total five inner compartments. 6. System should have high grade stainless steel interior and tough, powder coated exterior finish constructed on steel gauge. 7. Ultra thin vacuum insulation panel and silicon gasket on the doors for tight sealing to avoid formation of ice crystals and to reduce temperature recovery time after door opening. 8. Audible and visible security alarms for deviation from set temperature, power failure, probe failure, system failure, clean filter and door open, etc. 9. Eye level display of temperature parameters should be there. 10.	EPPENDORF INDIA PVT LTD	EPPENDORF, Model- U410	EACH	1,334,000.00	910,000.00	Kanada Agencies	11

21/08/19
 21/08/19
 21/08/19
 21/08/19

21/08/19

21/08/19


		4	5	6	7	8	8	9
100	<p>1. Product must have ISO 9001 safety requirements and H-C 61010 Electrical safety CE certified. 11. At least TWO years onsite warranty for complete equipment and FIVE years warranty on compressor should be provided. 12. Freezer should be supplied with 3KV compatible Servo voltage stabilizer with TIME DELAY FACTORY to restart after power failure. 13. INDIAN USERS LIST with contact address and telephone no. should be provided. 14. Technical features described above should exclusively be supported by authentic original illustrated company catalog of the instrument/equipment quoted. 15. Vendor's business record is to be authenticated by producing Income Tax Return for at least 3 previous years. 16. For smooth functioning, one 1.5 ton split Air conditioner to maintain ambient temperature is required with this deep freezer.</p>							
101	De humidifier	General	P.L. LANDON & CO	TANCO	CAPACITY 1.5 Jar /Each	149,200.00	98,000.00	Kanala Agencies
102	Digital Camera	Digital camera with high resolution minimum 13 Megapixels	NIKON	NIKON	EACH	176,000.00	92,000.00	Kanala Agencies
103	Digital Color Doppler System	The system should be latest fully Digital Color Doppler Ultrasound System and can be used for applications like Abdominal, OBS & Vascular Applications of Big and Small Animals. The system should have following essential features:	IMV INDIA PVT LTD	IMV, Model EXAGO(Made in France)	EACH	2,785,000.00	1,700,000.00	Kanala Agencies


21/06/2019
 HH-50/2019
 20/06/2019


27/08/2019

27/08/2019


1	2	3	4	5	6	7	8	9
103	Dog Ovulation Detector		DRAMNISKI	DRAMNISKI	EACH	95,000.00	57,000.00	Kamala Agencies L.I
104	EKG MACHINE	EKG 12 LEAD	ALLENCHERS MEDICAL SYSTEM PVT LTD	ALLENCHERS	EACH	1,240,200.00	648,000.00	Kamala Agencies L.I
105	Electrical dehormer	Electric Dehorming Instrument set for calves in age group 4-6 weeks with 1 tip of 17mm dia, 4 meter cable 24/170 watt transformer 230V for secure fast & bloodless dehorming in 10-15 seconds.	NATIONAL MEDITEK	NATOMED	EACH	3,610.00	2,350.00	Kamala Agencies L.I
106	Electrical sterilizer	Size: 300 x 150 x 125mm (12" x 6" x 5") Instrument Sterilizers, electric, with hinged lid fitted with immersion rod and water outlet tap, seamless stainless steel construction. Fitted with automatic cutoff safety pin device, thermostat & tray lifting arrangement.	NATIONAL MEDITEK	NATOMED	EACH	9,350.00	5,500.00	Kamala Agencies L.I

 21/06/19
 21/06/19
 21/06/19
 21/06/19

21.06.2019


21/06/19



1	2	3	4	5	6	7	8	9	
107	Electronic thawing unit	<p>The Thawing Unit - digital for thawing frozen semen straws should:</p> <ol style="list-style-type: none">1. Be portable with suitable handle2. Have suitable base with digital display and controls3. Have cylindrical vessel made of good quality stainless steel mounted on the base.4. Have concealed heating arrangement to provide uniform temperature throughout the water column.5. Have a tight lid.6. Dimension of Thawing Unit<ol style="list-style-type: none">a. Height/depth (inner) : 165 ± 5 mmb. Diameter (inner) : 65 ± 5 mmc. Digital display : indicate temperature in ° Celsius; should display the set temperature and the actual temperature of the water.7. The unit should maintain a pre-set temperature of 37° C (± 0.1° C)8. Have thermostatic control/electronic control of temperature with automatic cut off.9. Operating voltage 220V Rated Frequency : 50 Hz10. With electric cord of a minimum length of 2 meters with plug to suit 5 amp /3pin sockets.11. The entire surface of the inner wall of the cylinder should be smooth.12. Detailed colour photographs of the units offered should be submitted along with the tender13. Arrangement may be made to demonstrate the unit at the place and time which will be intimated by the evaluation committee.	IMV INDIA PVT LTD	IMV	EACH	26,900.00	18,200.00	Kamala Agencies	L1
108	Empty French mini straw (0.25ml)for integrated filling and sealing machine	PVC straw of different colours, pack of 2000 straws(0.25ml)	IMV INDIA PVT LTD	TBS FRENCH MINISTRAW	EACH/2000 Nos Pack	6,100.00	3,850.00	Kamala Agencies	L1



 11/08/2020

 11/08/2020

 11/08/2020



 11/08/2020






 11/08/2020



 11/08/2020

1	2	3	4	5	6	7	8	9
108	Empty French mini straw (0.25ml) for integrated filling and sealing machine	PVC straw of different colours. pack of 2000 straws(0.25ml)	IMV INDIA PVT LTD	ALPHA TBS FRENCH MINI STRAW	EACH/2000 Nos Pack	8,900.00	6,500.00	Kan Ager
109	EYE HOOK(LARGE)	Stainless steel with blunt ends	NATIONAL MEDITEK	NATOMED	EACH	380.00	250.00	Kan Ager
110	EYE HOOK(SMALL)	Stainless steel with blunt ends	NATIONAL MEDITEK	NATOMED	EACH	370.00	240.00	Kan Ager
111	Ear tag for Barcode Reading	SMILE compatibility, IMV make	IMV INDIA PVT LTD	IMV	EACH	3,350.00	1,980.00	Kan Ager
112	Ear tag	Ear tag for small animal	ALLFLEX INDIA PVT LTD	ALLFLEX	EACH	12.00	7.00	Kan Ager
112	Ear tag	Ear tag for Large Animal	ALLFLEX INDIA PVT LTD	ALLFLEX	EACH	14.00	8.00	Kan Ager
112	Ear tag	Ear tag applicator for Large Animal	ALLFLEX INDIA PVT LTD	ALLFLEX	EACH	900.00	500.00	Kan Ager
112	Ear tag	Ear tag applicator for small Animal	ALLFLEX INDIA PVT LTD	ALLFLEX	EACH	900.00	500.00	Kan Ager
113	Embryotome and Accessories	Thygesen model, nickel-plated with handle for fixation of obstetric chain.	NATIONAL MEDITEK	NATOMED	EACH	23,740.00	16,200.00	Kan Ager
114	Equipments for goat and sheep artificial instruments fresh and frozen semen	Consisting of Speculum- stainless steel, size 6" Goat AI pipette Test Tube- Borosilicate glass, size (ODXL) in mm=12X75 Mug-stainless steel, 500ml.	IMV INDIA PVT LTD	IMV	EACH	133,000.00	75,000.00	Ka Ag
114	Estrous Detector (for Cows)	1 Drminiski ESC Estrous Detector; 2 -1*9V 6F-22 Type Voating; 3- Manual	DRAMNISKI	DRAMNISKI	EACH	92,500.00	49,000.00	Ka. Ager

1	2	3	4	5	6	7	8	9
114	Estrous Detector(For Sheep and Goats)	1. Draminski LSC Estrous Detector; 2-1*9V 6F-22 Type Battery 3- Manual	DRAMINSKI	DRAMINSKI	EACH	92,500.00	49,000.00	Kamala Agencies LI
115	Elisa reader	1. System must be handy and user friendly. Instrument should be able to work stand-alone also. 2. System equipped with automatic plate loading and variable speed for plate shaking. 3. Automatic calibration before each reading 4. Compatible with various micro-plate platforms (FLAT, U, V-bottomed micro plate or up to 12 well strip plates) 5. Quick reading ability, single and /or dual wavelength selection 6. Self-diagnostic capabilities to detect lamp burnout at the start. 7. System must include compatible software for analysis and method performance validation. 8. Works between the wavelength ranges of 400 nm to 750 nm. 9. Photometric range between 0.0 to 3.0 OD and linearity range at least from 0.0 to 2.0 OD 10. System must provide accuracy not less than ± 0.01 for 0.0-3.0 OD at different filters and precision of 0.005 OD (for 0.0-3.0 OD) 11. High resolution of 0.001 OD 12. Up to 6 Filters slots with filters (e.g. 415,450,490, 500,655 nm) 13. Fast recording time not more than 10 sec 14. Data output USB interface with PC on board thermal printer (optional) 15. Date storage- Calendar/clock function 16. Provided with computer (RAM-4GB, Hard-drive-1 TB, Screen size-23" LED, CORE I3 fourth generation at least) for data storage and analysis 17. Software of High throughput, which can run a number assays in a single plate 18. Multiple plates processing with automatic data export. 19. Custom reporting function with performance verification and kinetics analysis 20. Provided with suitable stabilizer 21. Onsite warranty for at least 1 year. 22. Unit should operate at 220-240v / 50-60 Hz. 23. User's list & original literature must be submitted	BIOTEX ELX	BIOTEX ELX	BIOTEX ELX 800	669,200.00	478,000.00	Kamala Agencies LI

Approved
 20/10/2019
 20/10/2019

Accepted
 20/10/2019

20/10/2019
 20/10/2019

1	2	3	4	5	6	7	8	9
116	Electro cautery unit with Cautery blade for ear notching		NATIONAL MEDITEK	NATOMED	EACH	71,000.00	35,000.00	Kamala Agencies 1.1
117	ETO STERILIZER	SEMI AUTOMATIC/AUTO MATIC ETO STERILIZER SMALL SIZE/LARGE SIZE WITH CATRIGES 4.5 CFT,5CFT,8CFT,124 FT PCI MAKE	PEST CONTROL OF INDIA	PEST CONTROL OF INDIA	2 CFT MANUAL EACH	715,000.00	434,400.00	Kamala Agencies 1.1
117	ETO STERILIZER	SEMI AUTOMATIC/AUTO MATIC ETO STERILIZER SMALL SIZE/LARGE SIZE WITH CATRIGES 4.5 CFT,5CFT,8CFT,24 FT PCI MAKE	PEST CONTROL OF INDIA	PEST CONTROL OF INDIA	8 CFT AUTOMATIC EACH	5,486,000.00	3,180,000.00	Kamala Agencies 1.1
117	ETO STERILIZER	SEMI AUTOMATIC/AUTO MATIC ETO STERILIZER SMALL SIZE/LARGE SIZE WITH CATRIGES 4.5 CFT,5CFT,8CFT,24 FT PCI MAKE	PEST CONTROL OF INDIA	PEST CONTROL OF INDIA	24 CFT MANUAL EACH	3,265,000.00	1,980,000.00	Kamala Agencies 1.1
118	FIRSTAID BOX	General	NATIONAL MEDITEK	NATOMED	EACH	2,400.00	1,560.00	Kamala Agencies 1.1

By *[Signature]* 21.08.2019
 2.17.01.19
[Signature]

21/8/19
 Aff. Secy
 20/10/2022
[Signature]


1	2	3	4	5	6	7	8	9
241	Refrigerator	350 Ltr Capacity	GODREJ & BOYCE MANUFACTURING CO LTD	GODREJ	350 Ltr Each	45,300.00	36,400.00	Kamala Agencies L1
		Stabilizer	Autotake	Autotake	Each	7,000.00	3,000.00	Kamala Agencies
241	Refrigerator	99 Ltr Capacity	GODREJ & BOYCE MANUFACTURING CO LTD	GODREJ	99 Ltr Each	11,750.00	9,460.00	Kamala Agencies L1
		Stabilizer	Autotake	Autotake	Each	7,000.00	3,000.00	Kamala Agencies
241	Refrigerator	485 Ltr Capacity	GODREJ & BOYCE MANUFACTURING CO LTD	GODREJ	485 Ltr Each	63,100.00	50,200.00	Kamala Agencies L1
		Stabilizer	Autotake	Autotake	Each	9,000.00	4,200.00	Kamala Agencies
241	Refrigerator	595 Ltr Capacity	GODREJ & BOYCE MANUFACTURING CO LTD	GODREJ	595 Ltr Each	74,300.00	59,000.00	Kamala Agencies L1
		Stabilizer	Autotake	Autotake	Each	9,000.00	4,200.00	Kamala Agencies
242	Rat tooth forcep Dissecting	Stainless steel, straight, of 6", 8" length	NATIONAL MEDITEK	NATOMED	6" EACH	250.00	160.00	Kamala Agencies L1
242	Rat tooth forcep Dissecting	Stainless steel, straight, of 6", 8" length	NATIONAL MEDITEK	NATOMED	8" EACH	300.00	230.00	Kamala Agencies L1


22/08/2019


11/11/19

20/08/2019

1	2	3	4	5	6	7	8	8	9
244	Refrigerators, laboratory	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Upright type refrigerator. 5) External structure made of steel, anticorrosion treated; internal structure in stainless steel. 6) Storage compartment with a door complete of locking system. 7) Mounted on stabilizing adjustable feet. 8) Microprocessor controlled equipment. 9) Low electrical consumption. 10) Automatic diagnostic self-test 11) Adjustable temperature range at least from +2° to +15°C with resolution not higher than $\pm 1^\circ\text{C}$. 12) Internal volume storage capacity not less than 300 liters. 13) Provided with internal light. 14) Ventilated refrigeration system, to grant the maximum temperature uniformity and stability. 15) Automatic defrost system. 16) Outside control panel for temperature setting and monitoring. 17) Video/audio alarm at least for temperature set limits overtaking (over-temperature and under-temperature) and open doors. 18) Refrigerant Ozone-saving and CFC-free. 19) Almost 3 removable internal shelves (at least 1 of the three being a grid shelf). Additional consumables and accessories to be included	P.L. TANDON & CO	TANCO	EACH/ MODEL PLT 143	236,250.00	135,000.00	Kamala Agencies	L1


 2011/08/19
 2011/08/19
 2011/08/19

21.08.2019



 21/8/19

1	2	3	4	5	6	7	8	9
245	Refrigerators, Laboratory with freezer	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Upright type refrigerator with an independent freezer compartment, one above the other. 5) External structure made of steel, anticorrosion treated; internal structure in stainless steel. 6) Storage compartment divided into 2 fully insulated completely independent storage compartments, each with a door complete of locking system. 7) Mounted on stabilizing adjustable feet. 8) Microprocessor controlled equipment. 9) Low electrical consumption. 10) Automatic diagnostic self-test 11) Adjustable temperature range at least from -2° to $+10^{\circ}\text{C}$. 12) Freezer temperature range approx from -10° to -20°C . 13) Total (freezer + refrigerator) internal volume storage capacity around 300 liters, with single compartment net capacity around 100 - 150 liters. 14) Provided with internal light. 15) Ventilated refrigeration system, to grant the maximum temperature uniformity and stability. 16) Automatic defrost system. 17) Outside control panel for temperature setting and	SM Scientific Instruments Pvt Ltd	JINDAL	Each	880,000.00	450,000.00	National Enterprises L1


Handwritten signature


Handwritten signature


Handwritten signature

Handwritten signature

1	2	3	4	5	6	7	8	8	9
246	Refrigerators, laboratory Vaccine Conservation	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 (type). 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Refrigerator for laboratory pharmacy mainly used to preserve vaccines. 5) Microprocessor controlled equipment. 6) Equipment specifically designed to maintain a stable temperature 7) Inner glass door for easy inside observation. 8) Automatic diagnostic self-test. 9) Refrigerator adjustable temperature range approximately from +2° to +8°C with resolution not higher than +/- 1°C. 10) Internal volume storage capacity not less than 200 liters. 11) Internal light. 12) Internal air re-circulation for temperature stability improvement. 13) Automatic defrost system. 14) Spark-free internal chamber equipment. 15) LCD display for temperature setting and monitoring. 16) Visual and audio alarms for temperature set limits overtaking (over-temperature and under-temperature). 17) Refrigerant Ozone-saving and CFC-free. 18) Almost n° 3 removable internal grid shelves and at least n° 1 storage drawer. 19) Keyed door-lock. 20) Paper chart weekly temperature	P.L. TANDON & CO	TANCO	EACH/ MODEL PLT66	290,000.00	180,000.00	Kamala Agencies	L.I.


 21/08/19
 21/08/2019
 21/08/2019


 21.08.2019


 21/08/19

1	2	3	4	5	6	7	8	8	9
248	Realtime PCR Machine with UPS	<p>SPECIFICATIONS: Real Time PCR System</p> <ul style="list-style-type: none"> • Should be a peltier based thermal cycler. ; • Able to process 32 sample reactions, in 0.2ml tubes or 8-tube strips • Reaction volume should be from 10µl - 100µl ; • Temperature control method should be Peltier based ; • Temperature Range of 40-99°C with resolution of 0.01°C ; • Should provide Temperature uniformity of 0.05°C (SD); • Temperature Accuracy :+ 0.25°C ; • Fastest Speeds of 5°C/s heating and 4°C/s cooling • Run time to complete PCR cycle should be < 40 minutes ; • LED based Fluorescence excitation source 500nm (blue LED) ; • With no optical moving parts, • Fluorescence detection by CMOS array (510 to 750nm) ; • Should allow side reading enabling to use small reaction volumes, maintaining higher optical quality and ; • Flexible Fluorescence channels upto 120 optical channels ; • Sensitivity of Single copy detection upto 7 targets. ; • Multiplexing capability of upto 7 targets. ; • Should provide factory calibrated dyes of SYBR Green I, Resolight, FAM, VIC, HEX, Yellow 555, Red 610, TexasRed, Cy5, CAL 540, CAL 560, CAL 590, CAL 610, CAL 635, JOE, Pulsar 650, Quasar 570, Quasar 705, ROX, TAMRA, TET ; • Should also support user chosen custom dyes 	Himedia Laboratories Pvt Ltd	HiMedia	Each	3,169,000.00	1,970,000.00	National Enterprises	L1

20/08/19
20/08/19

28.08.2019

20/08/19

20/08/19

1	2	3	4	5	6	7	8	8	9
248		<ul style="list-style-type: none"> • System should support assay formats of Intercalating dyes (e.g. SYBR Green I), Hydrolysis Probes, Molecular Beacons, SimpleProbes, HybProbes ; • Operation should be through computer ; • Automated analysis module available for Absolute Quantification, Relative Quantification, Melting point Analysis, TaqMan Genotyping, HRM Genotyping ; • Should provide suitable computer with 500GB HDD, 2 GB RAM, USB ports, 1 network port, and windows 8 operating software ; • Interface or Connection via LAN, Direct connection to computer (RJ45), PC-free (USB stick) ; • Equipment should be of small foot print, light weight, quite operation ;with dimensions of W 25cm x D 27cm x H 23cm ; • Operating noise should be less than 40 db(A) ; • Electrical supply Voltage of 100 - 240V AC, 50-60Hz ; • Low power consumption of Power 170W ; • Should carry warranty of 1 year ; • Supplier should provide installation, demonstration and training at site. 							

Signature

2017/8/11

21.08.2019

2017/11/19
2017/11/19
2017/11/19

1	2	3	4	5	6	7	8	9	
249	Rumenotomy Set	Weingartpattern fixation frame for rumen made of non magnetic anti rust material having length : 27 cms. and breadth 18 cms. 1 Pc. 2. Weingartpattern fixation hook with 7 cm long made of anti rust material 6 Pc.; 3. Weignart pattern vulsellumforceps 24 cms made of stainless steel material 2 Pc.; 4. Mayo Pattern Scissor both end blunt straight blades made of stainless steel material screw joint having total length 16 cms with comfort finger loops. 1 Pc.; 5.do.....curved blades made of stainless steel material screw joint having total length 16 cms with comfort finger loops. 1 Pc.; 6. B.P. Handle Commercial No.3 made of stainless steel material and should take Blade No.11 firmly without any play 1 Pc.; 7.do.....4 made of stainless steel material and should take Blade No. 23 firmly without any play 1 Pc.; 8. BP Blades No. 11 and 23 ISI Marked 100 Pc.EachSterilised and Disposable; 9. Tissue forcejointless with 2x1 teeth stainless steel having total length 15 cms. 1 Pc.; 10. Allis Tissu Forceps screw joint 4x5 teeth stainless steel having total length 15 cms 2 Pc. ; 11. Backhaus Towel Clamp screw joint 14 cms length made of SS 4 Pc. ; 12. Halstead Mosquito Forcep with screw joints jaws having antiskid ; serrations straight , made of stainless steel total length : 13 cms 4 Pc. ;	NATIONAL MEDITEK	NATOMED	EACH	48,900.00	21,490.00	Kanala Agencies	1.1

Handwritten signature and date: 12/8/19

Handwritten signature and date: 12/8/19

Handwritten signature and date: 12/8/19

Handwritten signature and date: 12/8/19

1	2	3	4	5	6	7	8	8	9
		<p>13. Spencerwell's Artery Forcep Screw Joint straight jaws with antiskid serrations 15 cms length stainless steel. 2Pc.; 14.</p> <p>.....do.....curved blades 2 Pc.; 15. Mathew Pattern Needle Holder made of stainless steel, having total length 17 cm and screw jointed. 1 Pc.; 16. GroveDirector stainless steel winged 14 cms length 1 Pc.; 17. Suture Needle Traumatic reverse cutting triangular straight made of Stainless steel material having length of 93 mm and diameter 1.6mm 12 Pc.; 18. Suture Needle Traumatic reverse cutting triangular 1/2 circle made of stainless steel material having length of 102 mm and diameter 1.6mm 12 Pc.; 19. Suture Needle Atraumatic straight made of stainless steel material having length 93mm and diameter 1.6mm 12 Pc.; 20. Suture Needle Atraumatic curved 1/2 circle made of stainless steel material having length 102mm and diameter 1.6 mm 12 Pc.; 21. Surgical Gloves ISI Marked Disposable /Sterile size 6 10 Pairs do.....size, 7 10 pairssize 7 1/2" 80 Pairs ;22. Surgical Gown(Latha Green Color 100x 60mm with sleeves) 2 Nos.; 23. Face Mask 10 x 5 cms with arrangement for knots Green Latha Make 2 Nos.; 24. Surgical shroud (75x105cms) of latha green color with 17x2 cm slit in centre 2 Nos.; 25. Surgical shrouds 75 x 105 cms of latha green color 2 Nos. ;</p>							

2202/11/17

12.08.2019

100

1	2	3	4	5	6	7	8	9
		Fully Microprocessor based with inbuilt Internal weight Leveling knob Solid metal body Overload protection up to 100 kg. Maximum Capacity : 820 gm. Readability : 10 mg. Repeatability : 10 mg. Linearity : 0.2 mg. Power Supply: AC 230V/50Hz Tare Range: Full Range Compact and ergonomic in design High performance MONOLOC weighing cell Brilliant Backlit display Built in Application: Piece Counting, Percent weighing, Check weighing, Dynamic weighing, Statistics, Formulation, Totalling, Free Factor & Recall ISO 9001/ISO 14001 Certified For Quality And Environmental Standard	SARTORIOUS INDIA PVT LTD	SARTORIOUS	QUINTIX 1102-10 IN	124,700.00	72,000.00	Kamala Agencies

12-08-2019

21/8/19

21/8/19

1	2	3	4	5	6	7	8	9	
257	Straw filling,sealing and printer machine	To Fill, Seal and Print around 12,500 Straws (0.25 ml) per hour, 4 Straws at a time. Printing with an exclusive 50 microns nozzle fitted on Inkjet Printer for permanent and legible printing. 128 condensed Bar Code, 10 Digit printing in 22mm space. Liquid Level Sensor to check the semen level during process and automatically stop the machine when semen is finished in the reservoir. Auto Stop in case of any malfunction of the machine. Software with Straw color Indicator. Should Work as a Printing System only without changing/ removing/ separating any Should Work as a Filling and Sealing Machine only without changing/ removing/ separating any hardware. Solvent Suction System optional should be available to keep the laboratory air clean and healthy. Easy Operation and Maintenance. The System software can communicate and regulate Ultrasonic Generator, Vacuum Pump, Jet Printer and Filling & Sealing Machine. Technical assistance through modem possible. Compact, user friendly and ergonomic. Automatic identification of bill by using bar code. Computer piloted (MS windows compatible DEDICATED hardware, SOFTWARE). Compatible with Software for Management of Integrated Laboratory Environment , Low noise and Easy to clean . All parts that require cleaning should be easily handled without tools. Standardization :- ISO 9001 : 2000 Power Supply: 230 V 50-60 Hz.	IMV INDIA PVT LTD	ISEVO MINI	EACH	11,808,000.00	7,380,000.00	Kanala Agencies	1.1

Handwritten signature

2018/11/19

25.08.2019

Handwritten signature and date 25/08/2019


1	2	3	4	5	6	7	8	9
263	Suturing Needle 2 (straight with cutting edge)	2(Straight)with cutting edge	Newtek Instruments	Newtek	Each	90,00	49,00	National Enterprises L1
264	Suturing Needle 2 (straight with cutting edge)	2(Curved)with cutting edge	Newtek Instruments	Newtek	Each	90,00	49,00	National Enterprises L1
265	safety goggles	moulded of soft flexible polyvinyl chloride covering eyes ,nose and have vents to reduce fogging . The front lenses is made in clear acetate and can be worn over regular eye glasses	SAKET SURGICALS PVT LTD	SAKET	EACH	480,00	300,00	Kamala Agencies L1
266	Stretcher trolley for small animal	(SS) - Size: 210 L x 56W x 82H cms. • Frame work made of SS tubular pipe of 31 mm. dia mounted on four 15 cms. Dia. Castors (2 with brakes). • Removable stretcher top made of S.S. Sheet pressed on tubular frame. Brackets on alternate ends for IV drip rod.	NATIONAL MEDITEK	NATOMED	EACH	33,200,00	19,500,00	Kamala Agencies L1
267	Small animal operating table	Small animal operating table with following specifications: Hydraulic small animal surgery operation table made of high quality non-corrosive stainless steel. Table must have provision for hydraulically adjustable minimum height of 100 cm or more. Table should have loading capacity of 100 kg or more. Table should have tilttable top. Warranty should be least 2 years. Table should be supplied with standard accessories (to be mentioned by the vendor).	NATIONAL MEDITEK	NATOMED	EACH	185,000,00	96,000,00	Kamala Agencies L1


(Signature)
 2022/12/20
 2022/12/20


2022/12/20
 2022/12/20


(Signature)
 2022/12/20

1	2	3	4	5	6	7	8	9	
268	SHAKER INCUBATOR	Specifications: 1. Bench top Orbital Shaker Incubator with temperature range at least from 5°C to 60° C 2. Volume (liters) should not be less than 200 litre (at least for 12 flask of 1 lit & 30 flask of 500 ml) 3. Refrigeration- Hermetically sealed compressor uses CFC-free refrigerant 4. Micro controller based temperature indicator with digital display and PT 100 sensor. 5. Temperature accuracy- ± 0.5°C preferably by microprocessor PID controlled 6. Wave speed: 30 rpm to 350 rpm or better 7. Shaking Platform size should not be less than 420 mm x 420 mm 8. Display: LED or Digital display for indication of speed, running time and temperature. 9. Audio visual alarm if the temperature deviates from the preset temperature 10. Thick CFC free PU/F Insulation 11. Unit should operate with 220-240 volts/ 50-60 Hz 12. Heavy duty latch with lock & key 13. ISO 9001: 2008 and CE certification 14. Documents: Calibration report, Installation Qualification, Operational Qualification and Performance Qualification. Essential Accessories: 1. It should be supplied with clamp/adaptors for flasks as follows... Atleast For: 2000 ml- 10 clamps 1000 ml - 20 clamps 500 ml - 20 clamps 250 ml- 30 clamps 100 ml - 30 clamps 2. It should be supplied with compatible stabilizer. 3. It should come with 2 years warranty Additional: 1. Original illustrated Catalogue in support of technical specifications. 2. User list for the quoted model along with contact details including telephone no. and E. mail.	EPPEENDORF INDIA PVT LTD	EPPEENDORF (E24/F24R)	EACH	3,459,000.00	2,120,000.00	Kamala Agencies	L1
269	Stands micropipetty	1) Constructed of sturdy and noncorrosive material. 2) Capacity: 3 pipettes	Himedia Laboratories Pvt Ltd	Himedia	Each	4,154.00	3,410.00	National Enterpriser	L1


 Alex
 22/06/19
 22/06/19


 J. S. S.
 22/06/19


 J. S. S.
 22/06/19


 J. S. S.
 22/06/19