

1	2	3	4	5	6	7	8
270	Super sensitive Examination Protector gloves	Our highest quality. Disposable arm length gloves, with neck strap to make it self-holding for shoulder and armpit coverage. Smooth, extremely strong and with ultra high sensitivity. For rectal exams, artificial insemination and obstetrics. 90 cm long (from fingertip to neck opening). 2.5mu, 100% polyethylene.	IMV INDIA PVT LTD	IMV	EACH	17.00	8.90 Kamala Agencies L1
271	Sterilize cotton swab with container	For bacteriological & virological sample collection	Himedia Laboratories Pvt Ltd	HiMedia	100 Nos	1,512.00	National Enterprises L1
272	Sterilizing Drum	Step foot pedal operated Diameter-21.5 cm,height 33.5 cm,Diameter of take out bucket-15.5 cm height of take out -24.5 cm.capacity of the take out bucket-5 liter(ss)	Newtek Instruments	Newtek	Each	2,560.00	National Enterprises L1
273	Spirit lamp	Stainless Steel	Himedia Laboratories Pvt Ltd	HiMedia	Each	2,077.00	National Enterprises L1
274	Saline Set	Saline set meant for intra venous infusion of fluids	NATIONAL MEDITEK	NATOMED	EACH	136.00	Kamala Agencies L1
275	Scalven Set	To facilitate intra venous injection of small animals/dogs, Size: 20, 22, 24 Gz	NATIONAL MEDITEK	NATOMED	EACH	136.00	Kamala Agencies L1
276	Small Incubator	for inoculation of eggs and it's incubation for vaccine production	DAYAL POULTRY	DAYAL	EACH	790,000.00	Kamala Agencies L1

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

	3	4	5	6	7	8	8	9	
277	Suction apparatus	<p>Powder coated M.S. C stands.</p> <ul style="list-style-type: none">Noise level of suction apparatus from is 50dB \pm 0.3 dB.Leakage current of suction units is less than 84uA.Electrical requirement 220-230V, 50Hz, 1 Phase.Ideal for MFP / medical / surgical procedures.Rotary Vane Type Vacuum pump with- 740 \pm 10mm Hg.Free air displacement 40-60 ltrs/ minHeavy duty HN-65 antistatic castors, 320 Watt, 1440 RPM, 0.50 H. P. Electric Motor.100mm vacuum gauge.Non collapsible Silicon tubing.2 x 2.0 Ltrs. Polycarbonate jars with overflow safety.Bacterial filter fitted on top.	NATIONAL MEDITEK	NATOMED	EACH	24,100.00	15,200.00	Kamala Agencies	L1

21/08/19

21/08/19

21/08/19

21.08.2019

21/08/19

1	2	3	4	5	6	7	8	9
278	Spectro Photo meter	<p>Compact UV VIS spectrometer to quantify undiluted nucleic acids at nanolitre volume (1.5 µl) and other biomolecule analysis using standard volumes.</p> <p>Optical system : 1. Absorption beam ; 2. Reference beam ; 3. Light Source Xenon flash lamp; Receiver CMOS photodiode array; Wavelength range 200 nm to 830 nm; Wavelength Selection Method-dependent, freely selectable; Spectral bandwidth ≤ 4 nm; Wavelength increment 1 nm; Systematic wavelength error ± 1 nm; Random wavelength error ≤ 0.5 nm; Photometric measuring range 0.0 to 3.0 A at 260 nm ; Photometric reading accuracy $\Delta A : 0.001$; Random photometric error ≤ 0.002 at $A = 0$, ≤ 0.005 (0.5%) at $A = 1$; Systematic photometric error ± 1 % at $A = 1$; Cuvette compatibility : 1. Plastic disposable for standard methods ; 2. Standard Quartz cuvettes for standard methods ; 3. Microliter cuvette for low volume measurements. Micro cuvette type : • Coating: Hydrophobic surface coating ; • Cuvette blank : $< 0.1 A @ 230nm$, $\leq 0.05 A @ 260nm$; • Minimum vol: 1.5µl, dsDNA, 3 µl, Protein ; • Wavelength range: 180nm- 2000nm ; • Material: Quartz ; • Layer thickness: 1mm ; • Eight path height: 8.5mm ; Methods • Absorbance with one or more wavelengths, scans ; • Nucleic acids, Proteins, OD 600, dye labeling ; • Evaluation via factor, standard and calibration curve ; • Dual wavelength with subtraction and division evaluation ; Method dependent evaluation • Absorbance, concentration via factor and standard ; • Concentration via standard series using ; • Linear regression, Nonlinear regression with 2nd and 3rd degree polynomials ; • Spline analysis ; • Linear interpolation (point to point evaluation) ; • Absorbance allocation via subtraction and division ; • Ratio 260/280, 260/230.</p>	EPPENDORF INDIA PVT LTD	EPPENDORF EACH	1,796,000.00	868,000.00	Kamala Agencies	L1

27/05/2020
 Affile-82/2020

17/08/2019

27/08/2019

• molar concentration and total yield for nucleic acids ;
 • Frequency of incorporation of Cyt3, Cyt5 dyes and labeling
 • Density ;
 • Spectral zoom and peak evaluation for scans
 • Display 5.7" VGA 11-1 display ; Interfaces USB master for
 USB stick; USB slave for connection to PC ; Serial RS-232
 for thermal printer Memory > 100 method programs on the
 instrument ;> 1000 results with data, evaluation results and
 used parameters ;Power supply 100 to 240 V ; 10% / 50 to
 60 Hz \pm 5 % ; Power consumption Approx. 15 W in the
 operating step
 Approx. 5 W with dimmed display ;Dimensions W x D x H :
 295 mm x 400 mm x 150 mm ;Warranty: 1 year
 ;Accessories: 200 pcs of Disposable cuvettes should be
 supplied ;

0202/69/2020
D. J. H. H.
21/8/20
D. J. H. H.

12.2.2019

7/21/20

279	Teat Surgery Instruments	Teat Surgery set containing all following items in one proof cover with zipper :- Furstenberg's teat siphon with round head and side openings 9 cms length Nickel plated of diameter 2mm along with stillet 2 Pc Do.....2.5mm dia 2 Pc. Do.....3.0 mm dia 1 Pc. 2. Hug's Teat Tumor extractor stainless steel 12.5 cms length x 4mm dia 2 Pc. 3. Teat Dilator with screw action 17 cms 1pc 4. Teat Slitter with screw and 2 blades Danish Pattern 19cms length 3mm dianickel plated 1 pc 5. Teat slitter Swiss Model 9.5cms length 4 mm dia 1 pc 6. Lichty Teat Knife with sharp point made of stainless steel to cut the extra growth in teat canal. Total length with handle 18 cm. Made of stainless steel material	Newtek Instruments	Newtek	Each	3,700.00	2,200.00	National Enterpriser s	L1
280	Teat Syphon	stainless steel self retreating adjustable 2mm outer diam.	Newtek Instruments	Newtek	Each	48.00	23.00	National Enterpriser s	L1
281	Tissue Culture Plate	6 well plate	Himedia Laboratories Pvt Ltd	HiMedia	10 * 10 No	9,912.00	8,260.00	National Enterpriser s	L1
282	Towel Clip	Towel clip ss. 8cm	Newtek Instruments	Newtek	Each	340.00	200.00	National Enterpriser s	L1

12/08/2019
 25.08.2019
 20/07/2020
 20/07/2020

1	2	3	4	5	6	7	8	9
283	Trolley with wheel for dressing /Medicines	(SS) Standard Features: • Size: 76 L x 51 W x 82 H cms • Frame work made of S.S tubular pipe of 31 mm dia. • Mounted on four 10 cms. Dia castors. • Two fixed type S.S. Heavy gauge shelves. • Supplied complete with bucket and basin.	Newtek Instruments	Newtek	Each	21,100.00	14,300.00	National Enterprises
284	Thaw monitor	Should be made up of liquid crystal temperature indicating (32 OC- 40 OC) device with triple laminated and water resistant	NATIONAL MEDITEK	NATOMED	EACH	540.00	350.00	Kamala Agencies
285	Thermometer	Waterproof, with alarm and memory. 32oC-43oC. More than 200 hours of continuous operation on one battery 1.5V. Supplied in a strong plastic case.	NATIONAL MEDITEK	NATOMED	EACH	275.00	190.00	Kamala Agencies
286	Tamponing forcep	Made of stainless steel, length-18 inch The tip should have grooves to grip.	Newtek Instruments	Newtek	Each	1,780.00	950.00	National Enterprises
287	Tweezer	Stainless steel	NATIONAL MEDITEK	NATOMED	EACH	470.00	305.00	Kamala Agencies
288	Thermometer	OVAL	NATIONAL MEDITEK	NATOMED	EACH	150.00	78.00	Kamala Agencies
289	tenaculum	5", Stainless Steel	Newtek Instruments	Newtek	Each	1,730.00	950.00	National Enterprises
290	Trocacanal	SS for large animal & Small animal	Newtek Instruments	Newtek	Large Animal/Each	710.00	430.00	National Enterprises
290	Trocacanal	SS for large animal & Small animal	Newtek Instruments	Newtek	Small Animal/Each	550.00	370.00	National Enterprises

Handwritten signature/initials

Handwritten signature/initials

Handwritten signature/initials

Handwritten signature/initials

Handwritten signature/initials

1	2	3	4	5	6	7	8	9
291	Trolley for cryo containers	for 50L tr. Capacity container	PRECISE VACUUM SYSTEM PVT LTD	PRECISE	EACH	12,180.00	8,700.00	Kamala Agencies L.I
292	Test tube holder	Aluminium with wooden handle	NATIONAL MEDITEK	NATOMED	EACH	350.00	250.00	Kamala Agencies L.I
293	Technical specifications for portable color doppler veterinary ultrasound imaging system	1. Portable ultrasound system with rechargeable battery for scanning in small and large animals. 2. The system should come with species specific (at least for cattle/buffalo, sheep, goat, dog and horse) extensive calculation packages (at least for veterinary Obstetrics and Gynecology) and integrated imaging and calculation package for veterinary use. 3. System should have Real time 2D, M-mode, Pulsed wave Doppler, Color flow imaging and hardware. 4. System should have dual live mode or equivalent and dual and/or quad split display 5. System should have minimum 10" High Resolution medical grade flat LCD/LED monitor. 6. System should have speckle reduction technology with noise and artifact reduction for improving tissue clarity and margin definition. 7. System should have a dynamic range of at least 170dB 8. System should have at least 2 active transducer ports 9. Multifrequency, multifocal transducers to be supplied with the machine: Convex Array 2-5MHz (± 1 MHz), Linear rectal Transducer 4-7 MHz (± 1 MHz). System should have a dedicated port for ECG and should display the ECG simultaneously with the ultrasonogram on the screen 10. System should have facility for pan and zoom on live and frozen image.	IMV INDIA PVT LTD	IMV	EACH	3,257,000.00	1,750,000.00	Kamala Agencies L.I

20/07/2020

20/07/2020

20/07/2020

20/07/2020

20/07/2020

1	2	3	4	5	6	7	8	8	9
293		11. PW Doppler analysis in Auto & Manual mode. Variable sample volume, Selectable wall filter, Angle correction 12. System should have inbuilt DVD/CD RW, minimum four USB 3.0 ports and LAN/ Ethernet port 13. In-built hard drive capacity of minimum 160 GB 14. System should be DICOM 3 Compatible with Life time DICOM Licence and upgradable for PACS. 15. System should be CE or FDA (US) approved 16. The supplier should have installed at least 10 ultrasound machines of the same make in and around Bareilly region. 17. Accessories to be supplied: B/W Thermal printer of Latest model (with CE or FDA mark). Free Software upgrade(s) during the period of Warranty, Probe Protection covers / probe connector cover for all probes, External Hard Disk of at least 1 TB capacity. 18. Warranty of 2 years for the machine should be included in the package. ITEM							
294	Tray and rack ,slide staining	(Operation characteristics 1)Stainless-steel tray,2-four-bar rack for holding approx 22 slide 3 grill for holding 10 cover glasses,4- Screws for leveling rack	NATIONAL MEDITEK	NATOMED	EACH	2,200,00	1,335.00	Kamala Agencies	L1
295	Tissue culture bottle	pp	Himedia Laboratories Pvt Ltd	HiMedia	10 No	1,470,00	860,00	National Enterprises	L1

18/08/19

21/08/19

18.08.2019

18/08/19

21/08/19

21/08/2020

1	2	3	4	5	6	7	8	9	
296	Urine Analyzer	<p>Strips based on reflectance principle.</p> <p>One button operation: Results printed automatically. Gives results within 70 seconds. The analyzer contains light-emitting diodes (LEDs) that emit light at various wavelengths. Reading is done electronically.</p> <p>Parameter: pH, LFT, U (leukocytes), NIT (nitrate), PRO (protein), GLU (glucose), KET (ketone), UBG (urobilinogen), BIL (bilirubin), BLD (blood/hemoglobin)</p> <p>Consumption: Operating: maximum 15 W Standby: 1.3 W</p> <p>System Description: Light source: 6 LEDs Wavelengths: 565 nm (green) 3x 610 nm (orange) 3x Reader head: 1 head with 6 LEDs Printer: thermal printer Display: liquid crystal display, two lines of 24 characters each Memory: 100 samples Date/time: integrated clock Optimum Operating Conditions: Temperature: 68degree-78degree F (20degree-26degree C) Relative humidity: 30%-60%</p>	INDEX LABORATORIES	INDEX VET TEST	EACH	315,700.00	185,000.00	Kamala Agencies	L1
297	U.V Light	Ultra Violet Light, 3 feet, 35 Watt. Complete set of dual tubes for germicidal effect	PHILIPS	PHILIPS	EACH	6,790.00	4,440.00	Kamala Agencies	L1

21/08/19
 Affected
 21/08/2020
 21/08/19
 21/08/19

1	2	3	4	5	6	7	8	8	9
298	UV-Visible Spectrophotometer	For micro volume mode 1. Minimum Sample Size: 0.5 µL 2. Pathlength: 0.5 mm (auto ranging to 0.03 mm) 3. Light Source: Pulsed Xenon flash lamp 4. Detector Type: 2048 element CCD 5. Wavelength Range: 190-840 nm 6. Wavelength Accuracy: 1 nm 7. Absorbance Precision: 0.001 AU or 1%, whichever is greater 8. Absorbance Accuracy: 2% at 0.75 AU at 260 nm 9. Absorbance Range: 0.04 – 500 (1 cm equivalent) 10. Detection limit: 0.10 mg/ml BSA; 2.0 ng/µLs DNA 11. Measure maximum concentrated solutions 12. Measurement Time:	EPPIENDORF INDIA PVT LTD	EPPIENDORF	EACH	1,444,500.00	885,000.00	Kamala Agencies	L1
299	Ultrasound System (Dext Top)	Adjustable 17" LCD monitor (Resolution : 1280 X 1024) with touch screen facility, Fold down and lock mechanism for transportation, Brightness and contrast adjustment	IMV INDIA PVT LTD	EXAPAD	EACH	4,269,800.00	2,908,000.00	Kamala Agencies	L1

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

1	2	3	4	5	6	7	8	8	9
300	vet lab station (IVLS) with printer	Increases medical capabilities Integrate results from other IDEXX Vet Lab® analyzers and SNAP® test results into a single report Exhibit prior results in context with current result for easy comparison Exhibit results in organ - specific order easy abnormalities detection Graph-Patient result for monitoring changes Find all historical patient results with easy access to integrated reports and data. Boost staff's productivity : Reduce data entry with autofilling of client and patient information. Save time by running the entire IDEXX VetLab® suite of in-house analyzers from one interface. Reduce errors caused by manual data input Improve your business management Gain compliance through better record keeping. Ensure every test result is associated with that patient's electronic record. Link to your practice management system to ensure you have a complete record of every patient, including an accurate client invoice with no missed charges. Exhibits results in grouping Exhibit results listed by organ groups or hematology groups to help you quickly identify abnormalities or trends. Hematology results are grouped by RBC's, WBC's and platelets Chemistry results are grouped by kidney and liver, and further into primary and secondary groups for each organ. Snap Test results are recorded and included with diagnostic report Provided with Dedicated printer with software, FDA Approved	IDEXX LABORATORY ES	IDEXX	EACH	680,000.00	440,000.00	Kamala Agencies	1.1
302	Vacutainer with needle & needle holder	10 ML CAPACITY	BD INDIA PVT LTD	BD INDIA	EACH	47.00	29.00	Kamala Agencies	1.1

Handwritten signature

21/5/19

By 21.5.2019

Handwritten signature and date 21/5/2019

303	Vaccination kit	<p>Autoclavable graduated Plastic syringe with no traffic luer lock nozzle 5ml 1 Pc.</p> <p>Luer Lock Repeater syringe with adjustable Dose in 1.0ml interval</p> <p>Capacity: 5ml. with tube connector/And Ascending Tube 1 Pc.</p> <p>Poultry Vaccinator capacity 1 ml. with tube connector and Ascending Tube 1 Pc.</p> <p>Reusable Hypodermic Needles ISI Marked with brass hub sizes : 16Gx $\frac{3}{4}$" , 16Gx 1.1/2" 10</p> <p>18G x $\frac{3}{4}$" and 18G x 1.1/2" and 22 G x $\frac{3}{4}$" 10</p> <p>Pc. Each size</p> <p>Container with lid for used needles and syringes 1 Pc</p> <p>Cheate's forcep for syringe /Needle 1 Pc.</p> <p>Non electric Syringe /Needle Autoclave 1 Pc.</p> <p>All items should be packed in suitable shoulder carrying bag</p>	NATIONAL MEDITEK	NATOMED	EACH	26,200.00	18,900.00	Kamala Agencies	L1
-----	-----------------	--	---------------------	---------	------	-----------	-----------	--------------------	----

22.08.2019

22/8/19

20/08/19

20/08/19

20/08/2020

21/8/19

22/8/19

3	2	1	2	1	2	1	2	1	2	1	2	1
304	Vortex mixer	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Device mixing effect obtained by vortexes creation. 5) Microprocessor controlled equipment. 6) Maintenance-free shaking motor. 7) Adjustable speed up to a maximum value not lower than 2,200 rpm. 8) Capability to shake tubes with diameters up to at least 30 mm. 9) At least continuous and "touch start" operating modes. 10) Capability to work with standard tubes, flasks, beakers and microplates (not specifically designed or sold by just one-single brand/producer). Additional consumables and accessories to be included in the bid price; quote each separately. 11) Specific adaptors, holders, racks and/or platforms for mixing at least the following laboratory fluids/media containers. a) Flasks. b) Beakers. c) Micro-plates. 12) Dust cover	P.L. LANDON & CO	TANCO	EACH	170,000.00	89,000.00	Kamala Agencies	L1			
305	VACCINATOR	DOSE: FROM 0.1 CC TO 5ML BODY: BRASS, SILVER CHROMIUM COMP. ACCESSORIES LIKE, EXTRA WASHER SET, NIDDLE & SPANER	NATIONAL MEDITEK	NATOMED	EACH	10,680.00	5,600.00	Kamala Agencies	L1			

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

1	2	3	4	5	6	7	8	9
306	Water bath	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Heated, single chamber water bath, ease made of good quality material with static spray coating. 5) Inner chamber and top cover made of stainless anti-corrosion material. 6) Inner chamber working capacity not less than 10 liters. 7) Display for monitoring of at least: temperature setting and actual temperature (optional) 8) Temperature working range not smaller than from RT + 5°C up to 99°C (with the cover in place) 9) Thermostatic function/control over at least the entire equipment working temperature range. 10) Temperature measurement/control accuracy not higher than +/- 1°C. 11) At least visual (light) indications for at least: equipment switched ON and thermostat working. 12) Water bath compatible cover 13) Easy accessible heating element for any repair service. Additional consumables and accessories to be included in the bid price, quote each separately: 14) At least n° 1 rack/tray made of stainless anticorrosion material for not less than n° 10 standard testing tubes. 15) At least n° 1 rack/tray made of stainless	SM Scientific Instruments Pvt Ltd	JINDAL	Each	91,700.00	35,000.00	National Enterprises
308	Wrapping paper	Rolls	NATIONAL MEDITEK	NATOMED	EACH	285.00	190.00	Kamala Agencies
309	wire saw	Stainless Steel, 1 meter length & 2 handle.	Newtek Instruments	Newtek	Each	1,180.00	680.00	National Enterprises

Approved by
 2/2/2020
 2/2/2020

2/2/2020
 2/2/2020

1	2	3	4	5	6	7	8	8	9
		The equipment should be capable of performing following diagnostics tests for Animal semen with complete range of ready-to-use assays for semen analysis by flow cytometry 1. Viability and concentration 2. Mitochondrial activity 3. Oxidation molecule D 4. Oxidation molecule H 5. Viability and acrosome integrity 6. Bacterial concentration Should have following technical specifications: Software functionality an Assays: Instrument QC - Guava cheak™ software for evaluation and tracking of instrument performance. Instrument cleaning - Quick clean and shutdown automated cleaning functions.							
119	Flow Cytometer	Multicolor - Guava Express Plus™ Detection Assays - Express pro, instead red, green, yellow detection channels. Semen analysis software - CytoSoft, IMV based software for sample analysis with predefined string per species and assays. Data - Easysoft, IMV based software for citing, calculation and report presentations. Operating Environment: Temperature - 60-95 degree F (16-35degree C) Humidity - 10%-90% relative non-condensing. Power- 100-240 VAC, 50/60 Hz, 1 amp maximum current; Dimensions and Weight: System Size - height: 13"(33cm), width: 16"(40.6cm), Depth: 22.5"(57.2cm) System Weight - 73lbs (32.4kg) Laptop size, open (subject to variation without notice) - height: 10" (25.5cm), width: 13.27"(33.7cm).	IMV INDIA PVT LTD	ESYCYTE	EACH	14,895,000.00	9,400,000.00	Kamala Agencies	L1



21/08/19


21/08/2019


21/08/2019


21/08/19

1	2	3	4	5	6	7	8	9
119		<p>U-bottom (falcon 353910 polystyrene) other 9=6-well V and Flatbottoms.; Data management: Computer (subject to variation without notice) - FCS (flow cytometry stanard) 3.0 files, resultsfile for import into spreadsheet programs and optional export to FCS v.2 format; Additional features: Innovative capillary "Flow cell" technology - permits vacuum collection of cells, and eliminates the need for sheath fluid (no consumption of sheath fluid) The specific shape of the capillary enables the measurement of each spermatozoon. Thanks to a flow cell that excites cell in a plane surface not in one point, no need for alignment nor focusing. "flow cell" technology minimizes the use of reagents. Portable - the units compact size allows easy transport from one site to another. Safe - thanks to "flow cell" technology, the machine does not need to run in continuous mode. Tests are performed in a sealed environment inside the machine, in the dark, the capillary is cleaned and dried between each analysis. Tailored for semen - features diagnostic assays to evaluate the quality of semen. 96 well automated platform - An integrated 96 well automated platform that allows a repeatable, precise in the dark. Analysis of a conventional 96 well micro plates and 10 tubes.</p>						
120	Forcep Specimen Holder	Stainless steel with blunt ends & serrated tips of 10" length	NATIONAL MEDITEK	NATOMED	EACIT	3,450.00	2,200.00	Karala Agencies LI
122	Forzen semen straw lifter	Stainless Steel, L-shaped. Designed to lift semen straws with vertical serrations made at tip (8"-10")	NATIONAL MEDITEK	NATOMED	EACH	520.00	355.00	Kamala Agencies LI

 20/08/2019
 20/08/2019

 20/08/2019

 20/08/2019

 20/08/2019

1	2	3	4	5	6	7	8	8	9
126	fogging Machine	<p>The fogging machine should have the following features:</p> <ul style="list-style-type: none"> - Light Weight , portable & extremely user- friendly. - Machine should be highly versatile & ideal equipment to dispense wide range of chemical solutions like disinfectants, fungicides, germicides etc. - Droplets ULV size 6-30 μ range. - It should be equipped with superior quality VAC motors, non rotating vortex design nozzle assembly & precision flow metering valves. - Can treat 5000 ft³ or larger space within few minutes. - It should have rugged design power head housing & 5 liter chemical solution tank from corrosion resistant HDPE plastic. 	P.L. TANDON & CO	TANCO	EACH	79,000.00	32,000.00	Kamala Agencies	L1
127	Gynaecology aprons	Aprons for gynaecological work, of full length made of synthetic material and provided with braided cotton ends for tying with pockets. Should be made up of glass with silver lining. Should not have scratches.	Newtek Instruments	Newtek	Each	920.00	410.00	National Enterprises	L1
128	Gum boots Full size (in pair)	Gumboots should be of lightweight, made of rubber / synthetic rubber with cloth lining inner side. Sizes-8,9, 10 Full size Height - 14" Gumboots should be of lightweight, made of rubber / synthetic rubber with cloth lining inner side. Sizes-9, 10 & 12 Full size Height - 14" with steel toe.	NATIONAL MEDITEK	NATOMED	EACH	1,285.00	720.00	Kamala Agencies	L1

11/08/19

25.08.2019

21/8/19

21/8/19

1	2	3	4	5	6	7	8	8	9
129	Gradient thermal cycler	Thermal cycler with gradient facility with temperature differential ranges varying between 1-24°C. 2. Easy programming with facility to run 2 programs simultaneously having LCD display. 3. Block option should be 48x48x0.2ml, 96 well & 384 well format. 4. Fast ramping, for heating rate 2-3 rate °C per sec and cooling 3-4 °C per second. 5. Temperature range should be 0-100°C with accuracy ± 0.2°C and gradient accuracy ± 0.2°C at end rows. 6. Rapid well to well temperature uniformity of ±0.20 to 0.5°C. 7. Can store up to 1000 programmes on board with programming options of semigraphical, text based automatic. 8. Capable of password protected security features. 9. Onsite warranty of at least one year. 10. Provided with 2 KVA constant voltage stabilizer. 11. Heated lid with maximum temperature of >100 °C. 12. Original catalogue with users lists for the same model with complete address, phone numbers and Emails. 13. Unit should work at 220-240V/50-60 Hz.	EPENDORF INDIA PVT LTD	MASTER CYCLE NEXUS GX2	EACH	1,112,500.00	680,000.00	Kamala Agencies	L1
130	Gynecological obstetric embryotomy set	The instruments constitute following specification 1.1 lead snare 2.guard's chisel 3.fetal extractor 4.hoeblade castrating knife 5.krey's hook or tongs williams chisel 6.fetotome 8.threader for fetotome wire 9. cruch repeller 10. detorsion rod 11. william's blunt pointed cutting hook 12. williams long blunt hook 13. handles for wire saw 14. mui's chain handle 15.moore's obstetrical chains 16.gibbon's chain handle	NATIONAL MEDITEK	NATOMED	EACH	15,850.00	9,250.00	Kamala Agencies	L1
131	GLUTAVAC TEST	for the diagnosis of acute inflammation (increased immunoglobulin concentration) in large animal practice. The GLUTAVAC test reagent is supplied in ready-to-use vacu tubes (10ml) into which whole blood has to be drawn (like a standard blood sample).	Newtek Instruments	Newtek	Each	15,380.00	9,400.00	National Enterprises	L1

21/11/2020
21/11/2020

21.08.2019

2.1/8/19

1	2	3	4	5	6	7	8	8	9
132	Gel DOC SYSTEM	<p>System should have Image resolution >4 mega pixels for resolving closely spaced bands on a gel or blot. System should have 4.6x4.6 µm pixel size and >3.0 orders of linear dynamic range. Can accommodate maximum gel size (20x20 cm). System should be completely automatic & user does not have to zoom, focus, adjust, aperture or select light source. System should be modular with different sample trays & flexible to image a wide variety of applications, including nucleic acid. System should have UV, white light and optional blue light. System should have stain free capability for stain free gels and blots. Sample trays should be customizable as per user and must get recognized automatically. System require only One emission filter to accommodate a large portfolio of detection methods: EtBr, SYBR[®]Green, SYBR[®]Safe, SYBR[®]Gold, Gel Green, GelRed, Fast Blast[™], SYPRO Ruby, Flamingo[™] Dye, Oriole[™], CY3, Rhodamine, Green fluorescent protein, Hoechst, Krypton, Silver Stain, Copper stain, zinc stain, Coomassie Brilliant Blue, Coomassie Flour orange and other spectrally similar stains, labels and dyes. Should have lens flat-fielding calibration for each sample tray to deliver image data that are always optimized and reproducible without imaging artefacts, providing superior image uniformity and quantitation. Supplied with suitable power constant voltage stabilizer. Onsite warranty for atleast One year. Unit should work at 220-240V/50-60 Hz.</p>	VILBER, FRANCE	VILBER, Model : E BOX CX 5TS	EACH	1,456,000.00	990,000.00	Kamala Agencies	L1

20/08/2019
21/08/2019
21/08/2019

21.08.2019

21/08/19

21/08/19




1	2	3	4	5	6	7	8	8	9
		Software. Software should have highest level of automation in hardware calibration, image optimization, capture and analysis. Should have automated workflow recorded in a protocol file from image capture to results, thus eliminates need for training. Should allow 100% repeatability of the work flow by any user and ensure optimized image data and analysis from a gel in a single uninterrupted. fast and completely reproducible work flow. Should have automated image capture driven by a selected gel or blot application. Should have one button accusation from image capture to results. Should generate publication ready image with one click export option. Should generate customizable report.							
133	GPS tracking device	For tracking of movement of vehicle, 1. Installation of GPS Device with inbuilt GPRS module & dual embedded sim along with required accessories. 2) Mobile App development for geotagging and geo fencing, 3) hosting and maintaining android/ ios based mobile application,	Add Technologies India Ltd	Relycon	Each	24,200.00	14,500.00	National Enterprises	LI
		AMC including GPRS charges per month			AMC including GPRS charges per month		1,500.00	National Enterprises	LI
134	Gynaecology Aprons	Aprons for Gynecological work, of full length made of synthetic material.	Newtek Instruments	Newtek	Each	940.00	455.00	National Enterprises	LI

25/08/2019

2.7/8/19

25/08/2019

1	2	3	4	5	6	7	8	9
135	Gloves latex	Size-6, 7 & 7.5	Himedia Laboratories Pvt Ltd	HiMedia	Small Size	103.00	85.00	National Enterprises L1
135	Gloves latex	Size-6, 7 & 7.5	Himedia Laboratories Pvt Ltd	HiMedia	Medium/Large	103.00	85.00	National Enterprises L1
136	Gauge and bandage	Gauze cloth 90cm x 16 mtr Bandage 10 cm x 3 mtr ; 15 cm x 3 mtr	CHEMIE INDIA	CHEMIE INDIA	90 cm x 16 mtr	520.00	240.00	Kamala Agencies L1
136	Gauge and bandage	Gauze cloth 90cm x 16 mtr Bandage 10 cm x 3 mtr ; 15 cm x 3 mtr	CHEMIE INDIA	CHEMIE INDIA	60 cm x 16 mtr	340.00	150.00	Kamala Agencies L1
136	Gauge and bandage	Gauze cloth 90cm x 16 mtr Bandage 10 cm x 3 mtr ; 15 cm x 3 mtr	CHEMIE INDIA	CHEMIE INDIA	Roller Bandage (15cm x 3m)	260.00	120.00	Kamala Agencies L1
136	Gauge and bandage	Gauze cloth 90cm x 16 mtr Bandage 10 cm x 3 mtr ; 15 cm x 3 mtr	CHEMIE INDIA	CHEMIE INDIA	Roller Bandage (10cm x 3m)	198.00	89.00	Kamala Agencies L1
137	Gun for Barcode Reader	SMILE compatibility, IMV make	IMV INDIA PVT LTD	IMV	EACH	624,100.00	400,080.00	Kamala Agencies L1

1	2	3	4	5	6	7	8	8	9
138	Hand Operator Chaff Cutter (ISI MARKED)	<p>Component Material Relevant IS Standard</p> <p>Frame Cast Iron 210 : 1993</p> <p>Cover Plate Mild Steel 2062 : 1992</p> <p>Feed roll Shaft Mild Steel 2062 : 1992</p> <p>Tie rod Mild Steel 2062 : 1992</p> <p>Fly Wheel Cast Iron 210 : 1993</p> <p>Pulley Cast Iron/Standard Steel 210 : 1993</p> <p>Leg Mild Steel 2062 : 1992</p> <p>Leg Support Mild Steel 2062 : 1992</p> <p>Feed Chute Mild Steel 2062 : 1992</p> <p>Blade 2 Blades -High Carbon alloy Steel IS 1886 : 2000</p> <p>Blade Cover Mild Steel 2062 : 1992</p> <p>1. Cutter Head - It shall consist of fly wheel and fitted with two blade.</p> <p>2. Feeding System :- The feeding system in the chaff cutter shall be of chute type. The minimum 900 mm dia and weighing should not be less than 24 kg.</p> <p>3. Fly Wheel:- Minimum 850 mm and 60 mm width.</p> <p>4. Bearing - Original one.</p> <p>5. Blade: Blade should be of high carbon alloy steel having hardness 40-50 HRC.</p> <p>6. Main frame - 55 X 55 X 3 angle.</p> <p>7. Output - 75 kg. - 110 kg. per Hr.</p>	NEHA TOKA PVT LTD	LANDRA TOKA	EACH	14,800.00	9,990.00	Kamala Agencies	L1
139	Haemoglobinometer	manual	Himedia Laboratories Pvt	HiMedia	1 Nos	1,546.00	1,350.00	National Enterpriser	L1

21/08/19

21/08/19

21.08.2019

21/08/19

1	2	3	4	5	6	7	8	9
140	Hematology Analyzer with printer	Operating Temperature-Idexx vet Autoread -20-32 C(68 degree-90 degree F),idexx vet,centriguge -20-32C(68-90)F	IDEXX LABORATORIES	IDEXX VET TEST	EACH	2,085,000.00	1,420,000.00	Kamala Agencies L1
142	Hospital surgical pack	1. Artery forceps 6" straight 2 2.do.....curved 2 3. Allis Tissue forceps 6" 2 4. Allis Tissue forceps 8" 2 5. BP handle No.3 2 6. BP handle No. 4 2 7. Towel clamp 6 8. Surgical scissors straight 6 " 2 9.do.....curved 6" 2 10 Rat Tooth Forceps 6" 2 11. Plain Forces 6" 2 12. Needle Holder 8" 2 13. Self Retaining Forceps 2 14. Double end Probes 2 14. Trocar and Cannula set of 4 2 15. Suture Needles No. 1,2,3,4. 2 Pkts of each size 16. Surgical Tray ISI marked 12 x 10" with cover 1 17. Surgical shroud with slit 6 feet x 6 feet 1 18. Surgical drapes 3 x 3' 6 19. Hand Towel 12 x 12' 2 20. Surgical Gown 2 21. Wrap Cloth 4 x 4 feet 2	NATIONAL MEDITEK	NATOMED	EACH	18,700.00	11,200.00	Kamala Agencies L1

11/18/19
 11/18/19
 11/18/19

11/18/19
 11/18/19

11/18/19
 11/18/19

1	2	3	4	5	6	7	8	9
143	High quality large animal dystocia kit containing all the instruments	The kit should contain high quality imported obstetrical instruments Krey hook (stainless steel 8"L) Ostertag's Blunt eye hook (stainless steel 2"L) Harms Obstetrical hook (4" L) both sharp and blunt Robert's blunt eye hook Calving strap with D ring (60" strap with two D rings and 30" strap with two D rings T-bar obstetrical handle Moore's obstetrical chain handle (stainless steel) Obstetrical forceps for pigs Whelping forceps for small breeds (stainless steel 8"L) Obstetrical forceps for all domestic animals Obstetrical wire , wire guide and wire handle A portable case	Newtek Instruments	Newtek	Each	35,000.00	21,200.00	National Enterprises L1
144	HIGH SPEED CENTRIFUGE MACHINE	One non-removable swing out rotor complete with 8 tube carriers and rubber cushion, tube carriers to accommodate 15ml plain tubes, built in 4/5 step speed regulator, electrically operated.	EPPENDORF INDIA PVT LTD	EPPENDORF - MODEL 5702	EACH	690,000.00	455,000.00	Kamala Agencies L1
145	Hoof treating chute		NATIONAL MEDITEK	NATOMED	EACH	32,400.00	18,000.00	Kamala Agencies L1
147	Hoof cutter	High quality for professional users. Double action.	Newtek Instruments	Newtek	Each	35,000.00	18,200.00	National Enterprises L1

21.08.2019

21.08.2019

21.08.2019

21.08.2019

21.08.2019

1	2	3	4	5	6	7	8	9	
149	Hot plate	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Magnetic hot-plate microprocessor controlled stirrer. 5) Maintenance-free and soft-start stirring motor. 6) Chemical resistant hot-plate made of ceramic glass or Aluminium. 7) Adjustable stirrer speed with maximum value not lower than 1000 rpm. 8) Temperature range up to not less than 300°C. 9) Hot-plate dimension not less than 140mm of diameter or 140x140 mm (if the plate is not circle shape). 10) Speed and temperature settings visualization/display. 11) Equipment with over-temperature protection. 12) Internal electronic components protected against corrosion. 13) Connection for contact thermometer. Thermometer 14) Contact thermometer with digital display and temperature measurement range up to not less than 350°C with an accuracy not higher than +/- 1°C. 15) Equipment Integrated on external stand with temperature probe and/or thermometer holder. 16) Stainless steel or glass coated temperature probe with cable	SM Scientific Instruments Pvt Ltd	JINDAL	Each	85,200.00	46,800.00	National Enterprises	L1
150	Hot Air Oven	Laboratory standards	P.L. TANDON & CO	TANCO	EACH	78,000.00	32,000.00	Kamala Agencies	L1
151	Heat Detector		DRAMNISKI	DRAMNISKI	EACH	92,500.00	49,000.00	Kamala Agencies	L1
152	Hatcher		DAYAL POULTRY	DAYAL	EACH	1,167,000.00	690,000.00	Kamala Agencies	L1

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

1	2	3	4	5	6	7	8	8	9
153	Hot air oven		P.L. TANDON & CO	NATOMED	EACH	78,000.00	32,000.00	Kamala Agencies	L1
154	Immersion heater	Small size	NATIONAL MEDITEK	NATOMED	EACH	1,490.00	770.00	Kamala Agencies	L1
155	Inhalation anesthetic machine for small animals	<p>Technical Specification</p> <ul style="list-style-type: none">• M.S. Structure- Power coated• Castors with break in front.• Stainless steel table top with upper shelf for monitor.• O2/ N2O standard two tube flow meter.• Glass bottle Halothane Vaporizer• O2/ N2O yoke mounted pressure gauges for cylinder pressure• Gas specific 'A' type yokes for O2 & N2O.• Basic Nitrous cut-off system with audible O2 failure warning device.• Non-return valve fitted at back bar to protect the vaporizer from back pressure.• Emergency O2 flush with pressure relief valve.• CO2 Absorber• Magill's/ Bain's/Jackson rees breathing circuit• O2 Mox regulator with conversion kit & combination; spanner• Seal bodak& cork washer• Approx: (H) 1500 x mm (W) x 475 mm x (725 mm With Yokes) (D) x 600 mm• 50 KGS. (Approx.) Without cylinder.• Dog Mask (3 No.), Cat Mask (3No.), as standard accessories.	SAKET SURGICALS PVT LTD	SAKET	saam 7 each	315,000.00	225,000.00	Kamala Agencies	L1

Handwritten signature

2018/11/19

28.08.2019

Handwritten signature and date 28/08/2019

1	2	3	4	5	6	7	8	9
157	Incinerator, Bacteria		SM Scientific Instruments Pvt Ltd	JINDAL	Each	87,260.00	46,800.00	National Enterprises L1
158	Incinerator with Chimney	<p>Description of function and use: portable easy moving compact incinerator with zero maintenance cost designed for quick and safe disposable for complete incinerator of animal waste generate from farms, slaughter house & Veterinary Hospitals .</p> <p>Technical specifications - Must be protable Model.2.2-should be Microprocessor controlled.2.3-Furnace type with Burning Refuse.2.4-Vrick lined Enclosures 2.5-Dual chamber system works on thermal destruction using controlled combustion chamber 2.6-Highly insulated chamber having high temperature safe dozing and exhaust system 2.7-should have chamber having high temperature safe dozing and exhaust system.2.7-should have safe treatment of gases through long stack with less emission.2.8-Residue ash> 3% of the total waste pollution.complete with stack & water scrubbers.Must be fitted on castor wheels for easy movability</p>	P.L. TANDON & CO	TANCO	Tanco Model PLT 244/ Each	528,600.00	354,000.00	Kamala Agencies L1

22/12/19

22/12/19

22.08.2019

22/12/19

22/12/19

1	2	3	4	5	6	7	8	8	9
159	Incubator	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. 4) Equipment provided with an adequate Uninterruptible Power Supply (UPS) system with not less than 30 minutes of back-up. Operational characteristics 5) Floor standing, microprocessor controlled incubator. 6) Internal volume approx. 150 liters 7) Temperature range: approx from RT + 5°C to 75°C 8) Temperature stability at 37°C not higher than +/- 0.5°C 9) Corrosion resistant, easy to clean stainless steel interior. 10) Provided with glass inner door 11) Complete with control panel for setting and visualizing main parameters. 12) Visible and audible alarms for at least: over-temperature and incomplete cycle 13) At least one internal height adjustable shelf. 14) Blow/Fan air re-circulation	SM Scientific Instruments Pvt Ltd	JINDAL	Each	189,200.00	117,000.00	National Enterprises	L1
160	Insemination Gloves	Full arm length (34") LDPE disposable gloves with five fingers. The thickness should be 25 microns and above the joints should be strong and non-leaky. The gloves should be smooth, soft and harmless to the mucus membrane of cow / buffalo and also to the inseminator's hand.	IMV INDIA PVT LTD	IMV	EACH	12.00	6.90	Kamala Agencies	L1

21/08/2019

21.08.2019

2018/11/19