




1	2	3	4	5	6	7	8	8	9
202	Microscope	Electrical characteristics 1) A 120V, 60Hz mono-phase electrical source with line connection plug NEMA 5-15 type. 2) Protections against over-voltage and over-current line conditions. 3) Compliance with applicable Haitian standards and regulations. Operational characteristics 4) Binocular standard optical system biological microscope 5) Observation methods: brightfield, darkfield, and phase contrast 6) Adjustable inter-pupil distance. 7) Ocular tubes with adjustable focus. 8) Wide-field eyepieces at least 10x and 15x. 9) At least the following plan achromatic objectives provided: 4x, 10x, 40x, up to 100x (oil immersion). 10) Mechanical revolving nose piece with not less than 4 slots for interchangeable objectives positioning. 11) Integrated illumination system with a 30 W halogen lamp with adjustable light intensity or a Light-Emitting Diode (LED). 12) Coaxial focusing system with macro and micro movements. 13) Rack-less mechanical rectangular stage with horizontal movements in both two, x and y, perpendicular directions 14) Stage provided with easy removable pincers. 15) Equipment provided with interface to attach a digital camera (optional). 16) At least one interface for printer and PC connection for data printing and transmission (optional). Additional consumables and accessories to be included in the bid price: 17) Dust cover and hard box. 18) At least n° 1 spare halogen bulb if LED is not provided 19) Complete with the necessary start up kit of consumables and reagents for initial testing, calibration and training activities	Magnus OptoSystems India Pvt Ltd.	Magnus	Each	182,000.00	92,000.00	National Enterprises	L1


 27/08/19
 26/08/2020

Accounts Officer-I
 Directorate of A. H. & V. S.
 Odisha, Cuttack

21.08.2019



 20/08/19


 20/08/19

1	2	3	4	5	6	7	8	8	9
203	Microscope slide warmer	Temperature range ambient temp. to 70Â° C i.e. below the melting point of wax. The temperature is thermostatically controlled. The body is made out of thick mild steel duly finished in white stoving enamel/powder coated paint with mat finished color combinations and top is made of highly polished Stainless Steel sheet..To work on 220V AC 50 Hz single phase. Length x Width: a) 30 x 25 cm b) 40 x 25 cm	P.L. TANDON & CO	TANCO	TANCO MODEL PLT 144	19,600.00	10,800.00	Kamala Agencies	L1
204	Marker pen		Himedia Laboratories Pvt Ltd	HiMedia	Each	200.00	135.00	National Enterprises	L1

21.08.2019

21/8/19

21/8/19

21/8/19

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

21/8/19

1	2	3	4	5	6	7	8	8	9
205	Micro pipettes	<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none"> • Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes) • Adjustment opening: Adjust your pipettes to a specific liquid and volume. • Control Button: Very low operating force, Clolour indicates pipette volume. • Ejector: Very low operating force, positioned for perfect ergonomics. • It should be IVD Conformation. • Volume Display: 4 Digits. Magnifying shape. • Piston System: Ultra light system made of Forton. • Quick Connection Clip: Remove lower part easily. • Fully Autoclavable • Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>	Eppendorf India Pvt Ltd	Eppendorf	each Volume range: 0.1– 2.5ul	45,800.00	26,900.00	Kamala Agencies	L1

[Handwritten signature]
20/08/2019

[Handwritten signature]
21.08.2019

[Handwritten signature]
20/08/2019

[Handwritten signature]
20/08/2019

Accounts Officer-I
Directorate of A. H. & V. S.
Cuttack

1	2	3	4	5	6	7	8	9
		<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none"> • Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes) • Adjustment opening: Adjust your pipettes to a specific liquid and volume. • Control Button: Very low operating force, Clolour indicates pipette volume. • Ejector: Very low operating force, positioned for perfect ergonomics. • It should be IVD Conformation. • Volume Display: 4 Digits. Magnifying shape. • Piston System: Ultra light system made of Forton. • Quick Connection Clip: Remove lower part easily. • Fully Autoclavable • Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>						
205	Micro pipettes		Eppendorf India Pvt Ltd	eppendorf	each Volume range: 0.5 – 10ul	45,800.00	26,900.00	Kamala Agencies L1

6/10/12

27.08.2019

1111

5/28/12

22/08/2020
20/08/2020
Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
205	Micro pipettes	<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none"> • Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes) • Adjustment opening: Adjust your pipettes to a specific liquid and volume. • Control Button: Very low operating force, Colour indicates pipette volume. • Ejector: Very low operating force, positioned for perfect ergonomics. • It should be IVD Conformation. • Volume Display: 4 Digits. Magnifying shape. • Piston System: Ultra light system made of Forton. • Quick Connection Clip: Remove lower part easily. • Fully Autoclavable • Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>	Eppendorf India Pvt Ltd	eppendorf	each Volume range: 2 – 20ul	45,800.00	26,900.00	Kamala Agencies L1

[Handwritten signature]

[Handwritten signature]
21/8/19

21.08.2019

21/8/19

[Handwritten signature]
21/8/20

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
		Specification for Research plus Pipettes						
		<ul style="list-style-type: none">• Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes)• Adjustment opening: Adjust your pipettes to a specific liquid and volume.• Control Button: Very low operating force, Clolour indicates pipette volume.• Ejector: Very low operating force, positioned for perfect ergonomics.• It should be IVD Conformation.• Volume Display: 4 Digits. Magnifying shape.• Piston System: Ultra light system made of Forton.• Quick Connection Clip: Remove lower part easily.• Fully Autoclavable• Warranty- 3 Years						
205	Micro pipittes	<p>Volume range:</p> <p>0.1 – 2.5ul</p> <p>0.5 – 10ul</p> <p>2 – 20ul</p> <p>10-100ul</p> <p>20 – 200ul</p> <p>100 – 1000ul</p> <p>0.5 – 5ml</p> <p>1 – 10 ml</p>	Eppendorf India Pvt Ltd	eppendorf	each Volume range: 10-100ul	45,800.00	26,900.00	Kamala Agencies L1

[Handwritten signature]

[Handwritten signature]
21/08/19

[Handwritten signature]
21.08.2019

[Handwritten signature]

[Handwritten signature]
21/08/2019

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
205	Micro pipettes	<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none"> • Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes) • Adjustment opening: Adjust your pipettes to a specific liquid and volume. • Control Button: Very low operating force, Colour indicates pipette volume. • Ejector: Very low operating force, positioned for perfect ergonomics. • It should be IVD Conformation. • Volume Display: 4 Digits. Magnifying shape. • Piston System: Ultra light system made of Forton. • Quick Connection Clip: Remove lower part easily. • Fully Autoclavable • Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>	Eppendorf India Pvt Ltd	eppendorf	each Volume range: 20 – 200ul	45,800.00	26,900.00	Kamala Agencies	L1

Handwritten signature

21/08/19

21.08.2019

21/08/19

21/08/19

21/08/19

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9	
205	Micro pipettes	<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none">• Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes)• Adjustment opening: Adjust your pipettes to a specific liquid and volume.• Control Button: Very low operating force, Clolour indicates pipette volume.• Ejector: Very low operating force, positioned for perfect ergonomics.• It should be IVD Conformation.• Volume Display: 4 Digits. Magnifying shape.• Piston System: Ultra light system made of Forton.• Quick Connection Clip: Remove lower part easily.• Fully Autoclavable• Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>	Eppendorf India Pvt Ltd	eppendorf	each Volume range: 100 – 1000ul	45,800.00	26,900.00	Kamala Agencies	L1

[Handwritten signature]

[Handwritten signature]
21/8/19

23.08.2019

21/8/19

[Handwritten signature]

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
205	Micro pipettes	<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none"> • Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes) • Adjustment opening: Adjust your pipettes to a specific liquid and volume. • Control Button: Very low operating force, Colour indicates pipette volume. • Ejector: Very low operating force, positioned for perfect ergonomics. • It should be IVD Conformation. • Volume Display: 4 Digits. Magnifying shape. • Piston System: Ultra light system made of Forton. • Quick Connection Clip: Remove lower part easily. • Fully Autoclavable • Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>	Eppendorf India Pvt Ltd	eppendorf	0.5 – 5ml	45,800.00	26,900.00	Kamala Agencies	L1

21/08/19

21.08.2019

21/8/19

21/8/19

21/8/19

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9	
205	Micro pipettes	<p>Specification for Research plus Pipettes</p> <ul style="list-style-type: none">• Spring Loaded Tip Cone: Improved ergonomics, Tight fit to the tip (not in 5ml and 10ml pipettes)• Adjustment opening: Adjust your pipettes to a specific liquid and volume.• Control Button: Very low operating force, Clolour indicates pipette volume.• Ejector: Very low operating force, positioned for perfect ergonomics.• It should be IVD Conformation.• Volume Display: 4 Digits. Magnifying shape.• Piston System: Ultra light system made of Forton.• Quick Connection Clip: Remove lower part easily.• Fully Autoclavable• Warranty- 3 Years <p>Volume range: 0.1 – 2.5ul 0.5 – 10ul 2 – 20ul 10-100ul 20 – 200ul 100 – 1000ul 0.5 – 5ml 1 – 10 ml</p>	Eppendorf India Pvt Ltd	eppendorf	each Volume range: 1 – 10 ml	45,800.00	26,900.00	Kamala Agencies	L1

21/11/19

21.08.2019

21/8/19

21/8/19

Accounts Officer-
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
206	Muffle furnace(inner size-150 x 150 x300mm)	<p>Special Features: Digital single set point temperature Ceramic wool insulation permits faster heat up, reducing energy consumption. Door safety switch stops power to heating elements when door is opened, Equipped with thermocouple break protection, that helps preventing thermocouple failure runaway condition.</p> <p>Technical Specification:</p> <p>Material of Construction: Light weight with ceramic wool insulation(instead of Brick insulation). The outer casting is made of double walled thick PCRC sheet, duly painted with attractive stove enamel. Heating elements are made of A-1 wire and backed by high temperature ceramic wool insulation, which avoids loss of energy. The temperature control unit consists of Energy regulator, fitted in front of the furnace with two pilot lamps. The apparatus is complete with one pyrometer, thermocouple, silver thermal fuse, and main lead with power plug. To work on 220/230 Volts AC or 400/440 Volts AC.</p>	SM Scientific Instruments Pvt Ltd	JINDAL	Each	62,500.00	38,000.00	National Enterprises	L1
207	Needle Holder	6"(Mayo Hegar)	NATIONAL MEDITEK	NATOMED	EACH	520.00	370.00	Kamala Agencies	L1
208	N 95 Mask	Pore size- 95 nanometer	NATIONAL MEDITEK	NATOMED	EACH	175.00	85.00	Kamala Agencies	L1

21/08/19

21/08/19

21/08/19

21/08/19

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
209	NON -absorbent cotton	500 gm	Himedia Laboratories Pvt Ltd	HiMedia	Each	525.00	430.00	National Enterprises L1
210	Non IBR Boiler	Vertical oil fired, steam output -600 kg per hour	THERMAX LTD	THERMAX	EACH	2,975,600.00	1,740,000.00	Kamala Agencies L1
211	Nylon suture	Coarse fiber, in link	NATIONAL MEDITEK	NATOMED	EACH	15.00	7.00	Kamala Agencies L1
212	Nose Clamp	With spring . Easy to attach and remove	Newtek Instruments	Newtek	Each	1,290.00	700.00	National Enterprises L1
213	Nose Clamp	Bull holder, 100 cm, American model.	Newtek Instruments	Newtek	Each	1,290.00	700.00	National Enterprises L1

Handwritten signature
20/10/19

Handwritten signature
20/10/19

Handwritten signature
29.10.2019






Handwritten signature
20/10/19

Handwritten signature
20/10/19

Handwritten signature
20/10/19

Accounts Officer-I
Directorate of A. H. & V. S
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
214	NIR(Near Infrared Analyser)	feed & forge Analyser using near Spectroscopy, 1-Direct measurement of feed and ingredients samples in ground or un ground from e-direct measurement of palletes. 2-routine analysis of standard parameter such as fet, protine , fiber starch moisture etc. upgradable to othee parmeter on site if required, fast and easy operation and result for several constituents available in one minute-3-in build PC with an intuitive touch screen interface for easy operation , 4-instruments should have Ip 65 degree protection , so as to with stand humidity, dust, vibration and temprature fluctuations , 5-build in user friendly ISI scan NOVA software for routine operation. 5-Diagnostic software, 6-Transferable calibration from ane instruments to another, 6-suitable for networking using LAN (LOCAL) OR WAN INTERNATE, 7-SAMPLE CUP LARGE WITH LID - 1 NO, 8-SAMPLE CUP SMALL WITH HOLDER LID 1 NO 9-UPGRADABLE WITH ATTACHMENTS FOR LIQUID ANALYSIS CHECK CELL KIT. 10-256 PIXEL INGAAS DIODE ARRY. 11-WAVELENGTH RANGE 1100-1650NM. 120 OPTICAL BANDWIDTH 10.44+/-0.5NM. 12-SPECTRAL RESOLUTION 0.5 NM/DATA POINT. 13-NUMBER OF DATA POINTS 1100. 14-ABSORBANCE RANGE UP T2 AU, 15-WAVELENGTH ACCURACY 0.5 NM 16-WAVELENGTH PRECISION 0.05 NM (STANDARD DEVIATION)	FOSS ANALYTICAL (I) PVT LTD	MODEL DS1650	EACH	7,526,000.00	5,600,000.00	Kamala Agencies	L1
215	Orbital shaker with digital speed	30 to 300 RPM. 16 mm shoking orbit & 335X335 mm cradle platform. Ultra Clear RO water Plant reserve osmosis plus electro deionization with production rate of 20-350 lit./hour	P.L. TANDON & CO	TANCO	TANCO MODEL PLT 144	210,000.00	138,000.00	Kamala Agencies	L1

1	2	3	4	5	6	7	8	9
216	Olsen-Hegar needle holder	6" with cutting edge	NATIONAL MEDITEK	NATOMED	EACH	1,110.00	700.00	Kamala Agencies L1
217	Obstetric Set	Consists of: 180210 2 obstetric chains 1m	Newtek Instruments	Newtek	Each	13,100.00	8,700.00	National Enterprises L1
218	Operation Theatre table	(manual) for small animal surgery. Flat Top- 152.5 x 56 cm (60 x 22 in) flat top is made of heavy-gauge stainless steel with rounded corners, resulting in no welding, seams or sharp edges. The table center is slightly elevated so that fluids drain to a shallow through around table edges. The drain outlet on the ends has a hook for a collecting pail which is offered as an accessory. Two tie-down attachment on each side.	NATIONAL MEDITEK	NATOMED	EACH	134,400.00	96,000.00	Kamala Agencies L1
219	Operation Theatre light with voltage stabilizer	20DM/4 light <ul style="list-style-type: none"> • 20" dia dome with 4 reflectors of 7" dia. • The dome is made of ABS having epoxy powder coating and can be rotated minimum of 160 degree C vertically and laterally. • Fitted with imported halogen bulbs 24 V-70W x 4. • Lux 40,000 + 10% at 90 cm. • Heavy duty step down transformer made from special core with H. P. P. S. one a lid in base sterilizable central handle. • Heat absorbing & colour correcting special filter glass. • Stand and arm made of stainless steel. Backup power Supply (Inverter) / UP.	NATIONAL MEDITEK	NATOMED	EACH	122,700.00	65,000.00	Kamala Agencies L1
220	OT Lines 36" x 36" for covering instruments & trolley	36" x 36" for covering instruments & trolley	NATIONAL MEDITEK	NATOMED	EACH	860.00	450.00	Kamala Agencies L1

[Handwritten signature]

[Handwritten signature]
21/8/19

[Handwritten signature]
21.08.2019

[Handwritten signature]
Accounts Officer
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
221	OT Lines 36" * 36" (windowed)	36" * 36" (windowed)	NATIONAL MEDITEK	NATOMED	EACH	895.00	460.00	L1 Kamala Agencies
222	Office Set	Consist of Two table with draw (should of made of compressed wood),Two Main Chair, three side chair, Pen stand, 5 No of Chair for waiting area	GODREJ & BOYCE MFG CO LTD	GODREJ	EACH	85,790.00	66,800.00	L1 Kamala Agencies
222	Office Set for LAC	LAC-Steel Almirah-5 feet, S Shape Chair 1, Moulded Chair 2, Table with one side drawer-1 no;	GODREJ & BOYCE MFG CO LTD	GODREJ	EACH	39,639.00	30,900.00	L1 Kamala Agencies
222	Office Set for VD	VD- (1) Steel Almirah- 5 feet*2 nos-Size(), (Revolving Chairs- 3 nos, Size(), Moulded Chair- 5 nos, Size and Make), Table 4 feet*2.5 feet-with one side drawer -2 nos(size and make), Table 5 feet * 4 feet - 1 no (size and make)	GODREJ & BOYCE MFG CO LTD	GODREJ	EACH	116,624.00	90,920.00	L1 Kamala Agencies

21/08/19

21/08/19

21/08/19

21.08.2019

21/08/19

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9	
223	PCR Machine	<p>Specification for Gradient PCR Machine</p> <p>1. PCR machine with Block for 96 x 0.2ml PCR Tube, 71 x 0.5 ml PCR Tube; 2. 1 - 20°C Gradient with 12 different temperatures</p> <p>3. Gradient Technology ensures ramp rates are identical in both gradient and normal operation; 4. Heating and cooling via Triple Circuit</p> <p>5. Temperature control range: from 4 °C to 99°C</p> <p>6. Temperature Control Mode: Fast, Standard and Safe</p> <p>7. Lid Temperature range: 37 – 110°C; 8. Block Temperature Accuracy: $\pm 0.2^{\circ}\text{C}$; 9. Block Homogeneity: 20°C to 72°C; $\leq \pm 0.3^{\circ}\text{C}$, 95°C; $\leq \pm 0.4^{\circ}\text{C}$</p> <p>10. Ramping rate: approx. 3°C/s (heating) : approx. 2°C/s (cooling) ; 11. Block should maintain below 20°C during the lid reaching to its set temp. Which can avoid the miss amplification.; 12. Administrator and user login with or without PIN for enhanced security ;13. Preprogrammed template for easy selection from 16 temperature protocols viz. 2 step PCR, 3 Step PCR, Gradient PCR, Long Range PCR, Low volume PCR, Incubation, Cycle sequencing, Touch down PCR, Hot Start PCR, Hot Start PCR manual, Large volume PCR, Nested cycles, reduced ramping.;</p>	Himedia Laboratories Pvt Ltd	HiMedia	Each	790,000.00	475,600.00	National Enterprises	L1

[Handwritten signature]

[Handwritten signature]
21/8/19

[Handwritten signature]
21.08.2019

21/8/19

21/8/19

[Handwritten signature]
21/8/2019

Accounts Officer-I
Directorate of A. H. & V.
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
223		14. Time or Temperature increment with cycles in PCR program; 15. Adjustable ramp rate from 0.1° C to 3.0° C to meet critical amplification conditions ; 16. Customized programming allows a maximum of 20 steps and 99 cycles ;17. Auto Restart facility with user defined time interval when power fails and resumes ;18. System memory of more than 100 User folders and more than 700 programs ;19. Two USB ports: for Protocol transfer, Self-test, USB, printer / mouse ;20. E-mail Notification ; 21. Power save Standby function ;22. Option to connect up to TWO Mastercycler nexus eco for ultimate throughput. ;23. Optional Self-test dongle to check functionality of all 6 peltier elements ;24. Interface: Ethernet, CAN in, CAN out ;25. Calibration according to NIST (USA), DKD/PTB (Germany) UKAS/NPL (UK), UL/CUL listed ;26. Warranty: 1 Years							
224	PH Meter.Bench Top	Digital pH meter with LED/LCD display,range 0-14 pH with 0.01 resolution,auto/manual calibration,temperature range 00-1000 C electrically operated with electrode and electrode holder.	P.L. TANDON & CO	TANCO	EACH	15,120.00	9,000.00	Kamala Agencies	L1

27/12/19

28.08.2019

28/12/19

27/12/19

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
225	pulse-Oxy meter	Technical Specification: Monitoring SPO2, TEMP, PR & PLETH 2.8" Color TFT LCD, handheld storage/Review of trend 3888 groups data save time interval settable Bluetooth helps in connectivity to PC for easy review, storage or printing Audio & visual alarm when parameters exceeded and sensor -off. SPO2 SATURATION LEVEL TONE MODULATION Built-in rechargeable li-ion battery upto 15 hours running time suitable for canine, bovine, feline, equine, ovine applications.	SAKET SURGICALS PVT LTD	SAKET	POSSIP HVET/ EACH	105,600.00	64,800.00	Kamala Agencies L1
226	Ultrasound	The machine should be premium dedicated veterinary software portable digital b/w imaging system designed for canine, feline, ovine, equine, bovine etc. -b/w video printer MODES: B, B/B, B/M, and must have PW (pulse Wave)	IMV INDIA PVT LTD	IMV	EACH	1,596,000.00	950,000.00	Kamala Agencies L1
227	plaster cutting saw	Plaster cutting saw 316 LMV Steel	NATIONAL MEDITEK	NATOMED	EACH	2,350.00	1,400.00	Kamala Agencies L1
228	probang metal	All Metal	NATIONAL MEDITEK	NATOMED	EACH	2,400.00	1,400.00	Kamala Agencies L1
229	Polythene disposable examination hand gloves	Easy wear, all size, pkt contains 1000 nos	Himedia Laboratories Pvt Ltd	HiMedia	10 x 100 nos	9,744.00	8,420.00	National Enterprises L1

21.08.2019

21/8/19

21/8/19

21/8/19

Accounts Officer-I
Directorate of A. H. & V. S
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
230	plastic goblets (of different size)	plastic goblets 10,13,20,35,65 mm dia 118mm long for capacity of 15,25,50,200,700 nos,French mini straws respectively,the goblets should be made from virgin plastic compound. The goblets should withstand repeated cooling to -196 degree centigrade and thawing to ambient temperature.the goblets should be compatible with the imported make straw counting and filling machine.	IMV INDIA PVT LTD	IMV	10 MM Each	35.00	18.00	Kamala Agencies	L1
230	plastic goblets (of different size)	plastic goblets 10,13,20,35,65 mm dia 118mm long for capacity of 15,25,50,200,700 nos,French mini straws respectively,the goblets should be made from virgin plastic compound. The goblets should withstand repeated cooling to -196 degree centigrade and thawing to ambient temperature.the goblets should be compatible with the imported make straw counting and filling machine.	IMV INDIA PVT LTD	IMV	13 MM Each	42.00	19.50	Kamala Agencies	L1
230	plastic goblets (of different size)	plastic goblets 10,13,20,35,65 mm dia 118mm long for capacity of 15,25,50,200,700 nos,French mini straws respectively,the goblets should be made from virgin plastic compound. The goblets should withstand repeated cooling to -196 degree centigrade and thawing to ambient temperature.the goblets should be compatible with the imported make straw counting and filling machine.	IMV INDIA PVT LTD	IMV	35 MM Each	150.00	75.00	Kamala Agencies	L1

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]
21.08.2019

[Handwritten signature]

[Handwritten signature]

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
230	plastic goblets (of different size)	plastic goblets 10,13,20,35,65 mm dia 118mm long for capacity of 15,25,50,200,700 nos,French mini straws respectively.the goblets should be made from virgin plastic compound. The goblets should withstand repeated cooling to -196 degree centigrade and thawing to ambient temperature.the goblets should be compatible with the imported make straw counting and filling machine.	IMV INDIA PVT LTD	IMV	65 MM Each	225.00	115.00	Kamala Agencies	L1
231	Polyglactin suture	PGA No.1-0,2-0	FINE SURGICALS	STAR	A No.1-0,	2,050.00	1,100.00	Kamala Agencies	L1
232	Polydioxanone suture	PDS No.1, 2-0	FINE SURGICALS	STAR	Pkt of 12	2,840.00	1,450.00	Kamala Agencies	L1
233	300 ml poly propylene terile vaccine bottle with rubber stopper & aluminium cap	Round bottom and square (enclosed)	HAWKISS POLYMER	HAWKISS	ROUND	25.00	16.00	Kamala Agencies	L1
233	300 ml poly propylene terile vaccine bottle with rubber stopper & aluminium cap	Round bottom and square (enclosed)	HAWKISS POLYMER	HAWKISS	SQUARE	25.00	16.00	Kamala Agencies	L1
234	100 ml poly propylene round sterile vaccine bottle with rubber stopper & aluminium Cap	Round bottom and square (enclosed)	HAWKISS POLYMER	HAWKISS	ROUND	21.00	13.00	Kamala Agencies	L1

22.08.2019

21/11/19

22/08/2019
21/11/19
Accounts Officer-I
Accounts of A. H. & V. S.
Directorate of A. H. & V. S.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
234	100 ml poly propylene round sterile vaccine bottle with rubber stopper & aluminium Cap	Round bottom and square (enclosed)	HAWKISS POLYMER	HAWKISS	SQUARE	21.00	13.00	Kamala Agencies LI
235	Post Mortem Kit	1)Amputation saw, stainless steel 2)wire saw with handle, stainless steel (Gigil wire saw) 3)one spare wire for wire saw 4)post-mortem knife 5)full hand gloves, size-6 & 7 6)post-mortem gloves, latex, size-6 & 7 7)Bone hammer 8)bone chisel 9)post - mortem scissor, stainless steel 10)scalpel, stainless steel 11) All's tissue forceps, stainless steel 15" 12)sharpening stone 13)probe, stainless steel 14)scissor, dressing, stainless steel 15)Bowel scissors, stainless steel 16)retractor four prong sharp 17)Retractor four prong blunt 18)Dissecting forceps, toothed, stainless steel 19)Bone holding forceps, stainless steel 20)Abscess knife, stainless steel 22)Bone file 23)Measuring tape 24)Vice with two adjustable jaws made of iron for cutting heads of animal.	NATIONAL MEDITEK	NATOMED	EACH	21,900.00	14,600.00	Kamala Agencies LI
237	PVC Cuvette for Bovine photometer	Compatibility with Bovine Photometer	IMV INDIA PVT LTD	IMV	EACH	17.00	9.00	Kamala Agencies LI

28.08.2019

28/08/19

28/08/2019

28/08/2019

Accountants Officer-I
Directorate of A. H. & V. S
Odisha, Cuttack

28/08/19

1	2	3	4	5	6	7	8	9	
238	portable X-Ray Machine	The machine should be atleast 50mA, 85 KV, box type, portable X-ray machine. 2. It should have full wave rectified x-ray generator with stationary anode x-ray tube as per BIS standards 3. The preferred radiography rating should be in the range: - 50mA at 50KV, 40mA at 65KV, 35mA at 75KV, 30mA at 85KV. 4. The solid state electronic timer should have timing range from 0.06 to 6 seconds. 5. X-ray tube should have stationary anode with 1.8x1.8 mm focus. 6. The machine should have built in over voltage protection and indication. 7. The machine should have light beam collimator for centering and radiation protection. 8. Power requirement: single-Phase 230 volts, 15 Amps, 50 Hz. 9. Line voltage compensation from 210-250 Volt 10. The machine should be supplied with a compatible sturdy stand with heavy duty wheels and measuring tape for film focus distance. 11. Warranty two years minimum.	ALLEGERS MEDICAL SYSTEM PVT LTD	ALLENGER S	EACH	875,000.00	480,000.00	Kamala Agencies	L1
239	PPE KIT	a) Dengri- Disposable, free size full sleeve, polypropylene material 90 GSM b) House gloves: Disposable, ISI mark, 10 Nos c) Shoe cover: Disposable with thick padded foot base d) Face mask with Hepafilter e) Disposable protective glasses	NATIONAL MEDITEK	NATOMED	EACH	1,100.00	690.00	Kamala Agencies	L1

[Handwritten signature]

[Handwritten signature]
21/8/19

21.08.2019

[Handwritten signature]

[Handwritten signature]
Accounts Officer-I
Directorate of A. H. & V. S
Odisha, Cuttack

1	2	3	4	5	6	7	8	8	9
		<p>The power driven chaff Cutter will be of same specification as above with output of 3 Qtls. To 8 Qtls. Per Hrs.</p> <p>The weight of the chaff cutter will be 85 Kg. to 120 Kg. minimum without motor.</p> <p>Other requirement for both the chaff cutter (A) & (B) will be as per the norm given below :-</p> <ol style="list-style-type: none"> 1. The Chaff Cutter will be of BIS standard. 2. The Chaff Cutter should be portable with Safety majors as per BIS standards. 3. The Chaff Cutter must fulfill all the safety requirement as per IS 15542 :2005 (Power Operated Chaff Cutter Safety requirements) for feeding chute, fly wheel locking device, front safety guard, feeding chute cover, blade cover etc. 4. 1HP to 2 HP single phase electric motor with starter . 5. The chaff cutter must be tested at any Govt. of India testing Institute or approved State Agriculture, University. 6. Fly wheel type cutting mechanism with Chute fed type feeding mechanism. 7. The Chaff Cutter should have at least one year warranty period with repair, maintenance and manufacturing defect after post warranty period. 8. Assurance on availability of fast moving spare parts. 9. In case of power failure it can be operated manually. 10. The manufacturer must supply detail specifications, literature, photographs including list of Authorized dealers. 	NEHA TOKA PVT LTD	LANDRA TOKA	EACH	85,200.00	45,000.00	Kamala Agencies	L1
240	Power driven Chaff Cutter (ISI MARKED)								
241	Refrigerator	290 Ltr Capacity	GODREJ & BOYCE MANUFACTU RING CO LTD	GODREJ	290 Ltr Each	29,950.00	24,100.00	Kamala Agencies	L1
		Stabilizer	Autotake	Autotake	Each	7,000.00	3,000.00	Kamala Agencies	

By 22.08.2019

22/8/19

Accounts Officer-I
Directorate of A. H. & V.
Odisha, Cuttack

1	2	3	4	5	6	7	8	9
310	Water Bath with ss body	SS Body, 15-20 Ltr. Capacity, electrical and thermostatically controlled to maintain 20-500C	IMV INDIA PVT LTD	IMV	EACH	1,025,000.00	690,000.00	Kamala Agencies L1
311	water Purification system	RO SYSTEM / ULTRA PURIFICATION SYSTEM	SARTORIUS INDIA PVT LTD	SARTORIUS / MODEL ARIUM COMFORT	EACH	1,840,000.00	1,195,000.00	Kamala Agencies L1

Att. ext. 21

Accounts Officer-I
Directorate of A. H. & V. S.
Odisha, Cuttack

Dr. Anjan Kumar Das,
AD (Planning), DAH & VS,
Odisha, Cuttack

Sr. Scientific Officer, EPM
Directorate, Bhubaneswar

Shri Ramesh Kumar Swain,
Accounts Officer, DAH & VS,
Odisha, Cuttack

Shri L. K. Muduli,
FA-cum-Addl. Secretary to Govt. of Odisha,
F&ARD Department, Bhubaneswar

Shri Ratnakar Rout (IAS),
Director, AH & VS,
Odisha, Cuttack